

Analytical Summary Report

Analysis Method: 9040C

Analyst By : jignesh

Parameter: pH

Supervisor Review By : rubina

Run Number: LB136815

Slope : 98.6

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W3178
BUFFER PH 7.00 GREEN 1PINT PK6	W3093
PH 10.01 BUFFER,COLOR CD 475ML	W3191
buffer solution pH 7 yellow	W3217
Buffer Solution, PH2 (500ml)	W3161
pH 12.00 Buffer	W3200

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.05].

True Value of CCV2 = 12.00 Control Limits[+/- 0.05].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.3	4.01	08/14/2025	09:33
2	CAL2	1	Water	NA	NA	20.3	7.00	08/14/2025	09:34
3	CAL3	1	Water	NA	NA	20.3	10.02	08/14/2025	09:35
4	ICV	1	Water	NA	NA	20.3	7.01	08/14/2025	09:40
5	CCV1	1	Water	NA	NA	20.3	2.01	08/14/2025	09:44
6	Q2849-01	1	Water	NA	NA	24.1	4.90	08/14/2025	09:50
7	Q2856-01	1	Water	NA	NA	24.3	7.66	08/14/2025	10:00
8	Q2856-02	1	Water	NA	NA	24.2	7.79	08/14/2025	10:01
9	Q2856-03	1	Water	NA	NA	24.2	7.57	08/14/2025	10:02
10	Q2857-01	1	Water	NA	NA	24.6	7.27	08/14/2025	10:10
11	Q2857-02	1	Water	NA	NA	24.7	6.60	08/14/2025	10:17
12	Q2864-02	1	Water	NA	NA	23.2	6.31	08/14/2025	10:25
13	Q2864-04	1	Water	NA	NA	22.9	6.52	08/14/2025	10:30
14	Q2864-04DUP	1	Water	NA	NA	23.1	6.55	08/14/2025	10:33
15	CCV2	1	Water	NA	NA	20.3	12.02	08/14/2025	10:35

WORKLIST(Hardcopy Internal Chain)

VB 136815

WorkList Name : PH W Q2856 WorkList ID : 191257 Department : Wet-Chemistry Date : 08-14-2025 08:11:33

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
Q2849-01	A BUR-25-0071	Water	pH	Cool 4 deg C	PSEG03	J22	08/13/2025	9040C
Q2856-01	B #452	Water	pH	Cool 4 deg C	PSEG03	J42	08/13/2025	9040C
Q2856-02	A #455	Water	pH	Cool 4 deg C	PSEG03	J42	08/13/2025	9040C
Q2856-03	B #456	Water	pH	Cool 4 deg C	PSEG03	J42	08/13/2025	9040C
Q2857-01	B 4023	Water	pH	Cool 4 deg C	PSEG03	J12	08/13/2025	9040C
Q2857-02	B 3919	Water	pH	Cool 4 deg C	PSEG03	J12	08/13/2025	9040C
Q2864-02	B LAW-25-122-126	Water	pH	Cool 4 deg C	PSEG03	J21	08/13/2025	9040C
Q2864-04	B LAW-25-113-121	Water	pH	Cool 4 deg C	PSEG03	J21	08/13/2025	9040C

Date/Time 08/14/25 08:15
 Raw Sample Received by: [Signature]
 Raw Sample Relinquished by: [Signature]

Date/Time 08/14/25 13:30
 Raw Sample Received by: [Signature]
 Raw Sample Relinquished by: [Signature]