

## LB136820 INSTRUMENT ID:CV1

Method: 7471B

Operator: Admin

Date of Analysis: 14 Aug 2025 12:31:42

| Type | Sample ID       | Extended ID | Conc.   | $\mu$ Abs. | Units | Type | Std CondMethod | Date                 |
|------|-----------------|-------------|---------|------------|-------|------|----------------|----------------------|
| S    | 0.0 - 1         | 50.0        | -       | -884       | PPB   | Std  | 0.00007471B    | 14 Aug 2025 13:33:06 |
| S    | 0.2 - 1         | 50.2        | -       | 130        | PPB   | Std  | 0.20007471B    | 14 Aug 2025 13:35:23 |
| S    | 2.5 - 1         | 52.5        | -       | 11543      | PPB   | Std  | 2.50007471B    | 14 Aug 2025 13:37:40 |
| S    | 5.0 - 1         | 55.0        | -       | 24101      | PPB   | Std  | 5.00007471B    | 14 Aug 2025 13:40:00 |
| S    | 7.5 - 1         | 57.5        | -       | 37466      | PPB   | Std  | 7.50007471B    | 14 Aug 2025 13:42:19 |
| S    | 10.0 - 1        | 510.0       | -       | 50268      | PPB   | Std  | 10.00007471B   | 14 Aug 2025 13:44:39 |
| U    | ICV16 - 1       | ICV16       | 3.9867  | 19346      | PPB   | SMPL | -7471B         | 14 Aug 2025 13:51:43 |
| U    | ICB16 - 1       | ICB16       | -0.0198 | -1158      | PPB   | SMPL | -7471B         | 14 Aug 2025 13:53:59 |
| U    | CCV47 - 1       | CCV47       | 5.0185  | 24626      | PPB   | SMPL | -7471B         | 14 Aug 2025 13:56:17 |
| U    | CCB47 - 1       | CCB47       | -0.0130 | -1123      | PPB   | SMPL | -7471B         | 14 Aug 2025 13:58:34 |
| U    | CRA - 1         | CRA         | 0.2363  | 153        | PPB   | SMPL | -7471B         | 14 Aug 2025 14:00:55 |
| U    | HighStd - 1     | HighStd     | 10.0850 | 50554      | PPB   | SMPL | -7471B         | 14 Aug 2025 14:03:10 |
| U    | ChkStd - 1      | ChkStd      | 7.4872  | 37260      | PPB   | SMPL | -7471B         | 14 Aug 2025 14:05:25 |
| U    | PB169248BL - 1  | PB169248BL  | -0.0501 | -1313      | PPB   | SMPL | -7471B         | 14 Aug 2025 14:07:45 |
| U    | PB169248BS - 1  | PB169248BS  | 3.4388  | 16542      | PPB   | SMPL | -7471B         | 14 Aug 2025 14:10:06 |
| U    | Q2853-01 - 1    | Tr-5-081325 | 0.4355  | 1172       | PPB   | SMPL | -7471B         | 14 Aug 2025 14:12:23 |
| U    | Q2855-01 - 1    | #454        | 0.0147  | -981       | PPB   | SMPL | -7471B         | 14 Aug 2025 14:14:44 |
| U    | Q2858-01 - 1    | VNJ 216     | 0.7696  | 2882       | PPB   | SMPL | -7471B         | 14 Aug 2025 14:17:04 |
| U    | Q2858-01MS - 1  | VNJ 216MS   | 4.1886  | 20379      | PPB   | SMPL | -7471B         | 14 Aug 2025 14:23:18 |
| U    | CCV48 - 1       | CCV48       | 4.8665  | 23848      | PPB   | SMPL | -7471B         | 14 Aug 2025 14:26:12 |
| U    | CCB48 - 1       | CCB48       | -0.0206 | -1162      | PPB   | SMPL | -7471B         | 14 Aug 2025 14:28:29 |
| U    | Q2858-01MSD - 1 | VNJ 216MSD  | 4.4422  | 21677      | PPB   | SMPL | -7471B         | 14 Aug 2025 14:30:50 |
| U    | Q2858-01LX5 - 1 |             | 0.0081  | -1015      | PPB   | SMPL | -7471B         | 14 Aug 2025 14:33:06 |
| U    | Q2858-01A - 1   |             | 3.8724  | 18761      | PPB   | SMPL | -7471B         | 14 Aug 2025 14:35:24 |
| U    | Q2858-01DUP - 1 |             | 0.7846  | 2959       | PPB   | SMPL | -7471B         | 14 Aug 2025 14:37:39 |
| U    | CCV49 - 1       | CCV49       | 4.9235  | 24140      | PPB   | SMPL | -7471B         | 14 Aug 2025 14:39:58 |
| U    | CCB49 - 1       | CCB49       | -0.0153 | -1135      | PPB   | SMPL | -7471B         | 14 Aug 2025 14:42:16 |


LB135820  
INSTRUMENT ID: CV<sub>2</sub>

7471B

Linear

$\mu$  Abs.:

50268



A= 0.0000e+000

B= 1.9541e-004

C= 2.0644e-001

Rho= 0.9999249

Accept=Accepted

| Std ID | Conc.  | Calc.  | Dev.   | Mean  | SD or %RSD | Rep 1 | Rep 2 | Rep 3 | Rep 4 | Rep 5 | % D |
|--------|--------|--------|--------|-------|------------|-------|-------|-------|-------|-------|-----|
| 0.0    | 0.000  | 0.034  | 0.034  | -884  | 0.000      | -884  |       |       |       |       |     |
| 0.2    | 0.200  | 0.232  | 0.032  | 130   | 0.0 %      | 130   |       |       |       |       |     |
| 2.5    | 2.500  | 2.462  | -0.038 | 11543 | 0.0 %      | 11543 |       |       |       |       |     |
| 5.0    | 5.000  | 4.916  | -0.084 | 24101 | 0.0 %      | 24101 |       |       |       |       |     |
| 7.5    | 7.500  | 7.527  | 0.027  | 37466 | 0.0 %      | 37466 |       |       |       |       |     |
| 10.0   | 10.000 | 10.029 | 0.029  | 50268 | 0.0 %      | 50268 |       |       |       |       |     |

0.0000