

Analytical Summary Report

Analysis Method: 1010B

Reviewed By: rubina

Parameter: Flash Point

Supervisor Review By: jignesh

Run Number: LB136834

Ambient Barometric Pressure (mmHg): 755.00

Thermometer ID: Flash Point

Barometric Scale ID: 0511064

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W3194

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	9	28.00	82.4	82.7	08/14/2025	11:00
2	Q2856-01		1	13	100.00	>212.0	>212.0	08/14/2025	11:50
3	Q2856-01DUP		1	13	100.00	>212.0	>212.0	08/14/2025	12:20
4	Q2856-03		1	15	100.00	>212.0	>212.0	08/14/2025	12:50
5	Q2857-01		1	14	100.00	>212.0	>212.0	08/14/2025	13:20
6	Q2857-02		1	15	100.00	>212.0	>212.0	08/14/2025	13:50
7	Q2864-02		1	12	100.00	>212.0	>212.0	08/14/2025	14:20
8	Q2864-04		1	15	100.00	>212.0	>212.0	08/14/2025	14:50

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06

WORKLIST(Hardcopy Internal Chain)

6136834

WorkList Name : fp-08-14 WorkList ID : 191270 Department : Wet-Chemistry Date : 08-14-2025 09:28:57

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
Q2856-01	#452	Water	Flash Point	Cool 4 deg C	PSEG03	J42	08/13/2025	1010B
Q2856-03	#456	Water	Flash Point	Cool 4 deg C	PSEG03	J42	08/13/2025	1010B
Q2857-01	4023	Water	Flash Point	Cool 4 deg C	PSEG03	J12	08/13/2025	1010B
Q2857-02	3919	Water	Flash Point	Cool 4 deg C	PSEG03	J12	08/13/2025	1010B
Q2864-02	LAW-25-122-126	Water	Flash Point	Cool 4 deg C	PSEG03	J21	08/13/2025	1010B
Q2864-04	LAW-25-113-121	Water	Flash Point	Cool 4 deg C	PSEG03	J21	08/13/2025	1010B

Date/Time 08/14/2025 10:50
Raw Sample Received by: RM cws
Raw Sample Relinquished by: RM cws

Date/Time 08/14/2025 15:30
Raw Sample Received by: RM cws
Raw Sample Relinquished by: RM cws