

Analytical Summary Report

Analysis Method: 1010B

Reviewed By: rubina

Parameter: Flash Point

Supervisor Review By: jignesh

Run Number: LB136854

Ambient Barometric Pressure (mmHg): 760.00

Thermometer ID: Flash Point

Barometric Scale ID: 0511064

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W3194

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	9	28.00	82.4	82.4	08/18/2025	08:20
2	Q2856-05		1	8	19.00	66.2	66.2	08/18/2025	08:50
3	Q2856-05DUP		1	8	19.00	66.2	66.2	08/18/2025	09:20

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06

WORKLIST(Hardcopy Internal Chain)

LB136854

WorkList Name : fp-8-18 WorkList ID : 191313 Department : Wet-Chemistry Date : 08-18-2025 07:57:21

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
Q2856-05	#455	Solid	Flash Point	Cool 4 deg C	PSEG03	--Sele	08/13/2025	1010B

Date/Time 08/18/2025 08:10
Raw Sample Received by: RM (wts)
Raw Sample Relinquished by: RM (wts)

Date/Time 08/18/2025 10:00
Raw Sample Received by: RM (wts)
Raw Sample Relinquished by: RM (wts)