

Analytical Summary Report

Analysis Method: 1010B

Reviewed By: rubina

Parameter: Flash Point

Supervisor Review By: jignesh

Run Number: LB136857

Ambient Barometric Pressure (mmHg): 760.00

Thermometer ID: Flash Point

Barometric Scale ID: 0511064

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W3194

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	9	28.00	82.4	82.4	08/18/2025	08:20
2	Q2890-01		1	14	100.00	>212.0	>212.0	08/18/2025	10:20
3	Q2890-01DUP		1	14	100.00	>212.0	>212.0	08/18/2025	10:50
4	Q2890-02		1	15	100.00	>212.0	>212.0	08/18/2025	11:20
5	Q2890-03		1	15	100.00	>212.0	>212.0	08/18/2025	11:50
6	Q2890-04		1	12	100.00	>212.0	>212.0	08/18/2025	12:20

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06

WORKLIST(Hardcopy Internal Chain)

6136857

WorkList Name : fp-08-18 WorkList ID : 191315 Department : Wet-Chemistry Date : 08-18-2025 07:58:11

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
Q2890-01	60296	Water	Flash Point	Cool 4 deg C	PSEG03	D21	08/15/2025	1010B
Q2890-02	60295	Water	Flash Point	Cool 4 deg C	PSEG03	D21	08/15/2025	1010B
Q2890-03	60272	Water	Flash Point	Cool 4 deg C	PSEG03	D21	08/15/2025	1010B
Q2890-04	60294	Water	Flash Point	Cool 4 deg C	PSEG03	D21	08/15/2025	1010B

Date/Time 08/18/2025 10:10
 Raw Sample Received by: RM (wcl)
 Raw Sample Relinquished by: RM (wcl)

Date/Time 08/18/2025 13:10
 Raw Sample Received by: RM (wcl)
 Raw Sample Relinquished by: RM (wcl)