


LB136858
INSTRUMENT ID: CV1

7471B

Linear

μ Abs.:

63133



A= 0.0000e+000

B= 1.5905e-004

C= 4.4238e-002

Rho= 0.9979941

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%
0.0	0.000	0.072	0.072	173	0.000	173					0
0.2	0.200	0.206	0.006	1020	0.0 %	1020					3
2.5	2.500	2.297	-0.203	14162	0.0 %	14162					8
5.0	5.000	5.389	0.389	33604	0.0 %	33604					8
7.5	7.500	7.150	-0.350	44678	0.0 %	44678					5
10.0	10.000	10.086	0.086	63133	0.0 %	63133					1

LB136858 INSTRUMENT ID:CV1

Method: 7471B

Operator: Admin

Date of Analysis: 18 Aug 2025 10:43:59

Type	Sample ID	Extended ID	Conc.	μ Abs	Units	Type	Std Conc	Method	Date
S	0.0 - 1	SC.0	-	173	PPB	Std	0.0000	7471B	18 Aug 2025 13:14:55
S	0.2 - 1	SO.2	-	2748	PPB	Std	0.2000	7471B	18 Aug 2025 13:17:12
S	2.5 - 1	S2.5	-	14162	PPB	Std	2.5000	7471B	18 Aug 2025 13:22:05
S	5.0 - 1	S5.0	-	33604	PPB	Std	5.0000	7471B	18 Aug 2025 13:24:22
S	7.5 - 1	S7.5	-	44678	PPB	Std	7.5000	7471B	18 Aug 2025 13:28:08
S	10.0 - 1	S10.0	-	63133	PPB	Std	10.0000	7471B	18 Aug 2025 13:30:24
U	ICV20 - 1	ICV20	4.1630	25896	PPB	SMPL	-7471B		18 Aug 2025 13:54:09
U	ICB20 - 1	ICB20	0.0069	-235	PPB	SMPL	-7471B		18 Aug 2025 13:58:24
U	CCV60 - 1	CCV60	4.5478	28315	PPB	SMPL	-7471B		18 Aug 2025 13:58:41
U	CCB60 - 1	CCB60	-0.0027	-295	PPB	SMPL	-7471B		18 Aug 2025 14:00:57
U	CRA - 1	CRA	0.2341	1194	PPB	SMPL	-7471B		18 Aug 2025 14:03:14
U	HighStd - 1	HighStd	10.1246	63378	PPB	SMPL	-7471B		18 Aug 2025 14:05:29
U	ChkStd - 1	ChkStd	7.0732	44193	PPB	SMPL	-7471B		18 Aug 2025 14:07:44
U	PB169281BL - 1	PB169281BL	-0.0132	-361	PPB	SMPL	-7471B		18 Aug 2025 14:10:04
U	PB169281BS - 1	PB169281BS	3.5709	22173	PPB	SMPL	-7471B		18 Aug 2025 14:12:23
U	Q2879-01 - 1	OU4-TS-GRILLO-TSCP03-081425	0.9052	5413	PPB	SMPL	-7471B		18 Aug 2025 14:14:39
U	Q2879-03 - 1	OU4-TS-GRILLO-TSCP04-081425	1.1164	6741	PPB	SMPL	-7471B		18 Aug 2025 14:16:57
U	Q2879-05 - 1	OU4-TS-GRILLO-TSCP05-081425	0.7498	4436	PPB	SMPL	-7471B		18 Aug 2025 14:19:17
U	Q2879-07 - 1	OU4-TS-GRILLO-TSCP06-081425	0.4705	2680	PPB	SMPL	-7471B		18 Aug 2025 14:21:36
U	Q2879-09 - 1	OU4-TS-GRILLO-TSCP07-081425	1.4929	9108	PPB	SMPL	-7471B		18 Aug 2025 14:23:51
U	CCV61 - 1	CCV61	5.0714	31607	PPB	SMPL	-7471B		18 Aug 2025 14:26:07
U	CCB61 - 1	CCB61	-0.0080	-318	PPB	SMPL	-7471B		18 Aug 2025 14:28:24
U	Q2879-11 - 1	OU4-TS-GRILLO-TSCP08-081425	1.9691	12102	PPB	SMPL	-7471B		18 Aug 2025 14:30:41
U	Q2879-13 - 1	OU4-TS-GRILLO-TSCP09-081425	1.3663	8312	PPB	SMPL	-7471B		18 Aug 2025 14:32:57
U	Q2879-15 - 1	OU4-TS-GRILLO-TSCP10-081425	0.7046	4152	PPB	SMPL	-7471B		18 Aug 2025 14:35:15
U	Q2879-17 - 1	OU4-TS-GRILLO-TSCP11-081425	0.4764	2717	PPB	SMPL	-7471B		18 Aug 2025 14:37:32
U	Q2883-01 - 1	RBR2520553	5.9169	36923	PPB	SMPL	-7471B		18 Aug 2025 14:39:48
U	Q2884-01 - 1	VNJ-232	1.3501	8210	PPB	SMPL	-7471B		18 Aug 2025 14:42:05
U	Q2884-03 - 1	302	0.4513	2558	PPB	SMPL	-7471B		18 Aug 2025 14:44:25
U	Q2888-01 - 1	VNJ-210	0.9974	5993	PPB	SMPL	-7471B		18 Aug 2025 14:46:43
U	Q2889-01 - 1	PL-02-081525-	0.4902	2804	PPB	SMPL	-7471B		18 Aug 2025 14:49:00
U	Q2889-01DUP - 1	PL-02-081525-DUP	0.6431	3765	PPB	SMPL	-7471B		18 Aug 2025 14:51:17
U	CCV62 - 1	CCV62	4.8304	30092	PPB	SMPL	-7471B		18 Aug 2025 14:53:32
U	CCB62 - 1	CCB62	0.0951	320	PPB	SMPL	-7471B		18 Aug 2025 14:55:48
U	Q2889-01MS - 1	PL-02-081525-MS	4.1160	25600	PPB	SMPL	-7471B		18 Aug 2025 14:58:06
U	Q2889-01MSD - 1	PL-02-081525-MSD	4.1403	25753	PPB	SMPL	-7471B		18 Aug 2025 15:00:22
U	Q2889-01LX5 - 1		0.0512	44	PPB	SMPL	-7471B		18 Aug 2025 15:02:39
U	Q2889-01A - 1		4.2124	26206	PPB	SMPL	-7471B		18 Aug 2025 15:04:58
U	Q2857-03 - 1		12.1389	76042	PPB	SMPL	-7471B		18 Aug 2025 15:07:14
U	Q2857-03DLX5 - 1		2.2395	13802	PPB	SMPL	-7471B		18 Aug 2025 15:18:29
U	CCV63 - 1	CCV63	4.6237	28792	PPB	SMPL	-7471B		18 Aug 2025 15:26:09
U	CCB63 - 1	CCB63	0.0345	-61	PPB	SMPL	-7471B		18 Aug 2025 15:31:20