

Analytical Summary Report

Analysis Method: 1010B

Reviewed By: rubina

Parameter: Flash Point

Supervisor Review By: jignesh

Run Number: LB136863

Ambient Barometric Pressure (mmHg): 760.00

Thermometer ID: Flash Point

Barometric Scale ID: 0511064

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W3194

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	9	28.00	82.4	82.4	08/18/2025	08:20
2	Q2895-01		1	14	100.00	>212.0	>212.0	08/18/2025	13:35
3	Q2895-01DUP		1	14	100.00	>212.0	>212.0	08/18/2025	14:05

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06

WORKLIST(Hardcopy Internal Chain)

6136863

WorkList Name : flash point-2895-

WorkList ID : 191333

Department : Wet-Chemistry

Date : 08-18-2025 11:10:32

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
Q2895-01	1104	Solid	Flash Point	Cool 4 deg C	PSEG03	D31	08/18/2025	1010B

Date/Time 08/18/2025 13:25
 Raw Sample Received by: RM (w)
 Raw Sample Relinquished by: RM (w)

Date/Time 08/18/2025 14:50
 Raw Sample Received by: RM (w)
 Raw Sample Relinquished by: RM (w)