

66136 896

Test results

Aquakem 7.2AQ1

Page: 1

Alliance Technical Group
284 Sheffield Street, Mountainside, NJ 07092

Reviewed by : RM Instrument ID : Konelab

8/20/2025 15:01

Test: Phenolics-

Sample Id	Result	Dil. 1 +	Response	Errors
ICV1	1.015	0.0	0.109	
ICB1	0.001	0.0	0.008	
CCV1	0.988	0.0	0.106	
CCB1	0.001	0.0	0.008	
PB169316BL	0.002	0.0	0.008	
PB169316BS	1.012	0.0	0.109	
Q2893-10	0.027	0.0	0.011	
Q2893-11	0.053	0.0	0.013	
CCV2	0.994	0.0	0.107	
CCB2	0.002	0.0	0.008	

N 10
Mean 0.410
SD 0.5106
CV% 124.66

Aquakem v. 7.2AQ1

Results from time period:

Wed Aug 20 14:49:06 2025

Wed Aug 20 14:55:59 2025

Sample Id	Sam/Ctr/c	Test short r	Test type	Result	Result unit	Result date and time	Stat
0.00PPM	A	Phenolics-	P	-0.0004	mg/l	8/20/2025 13:37:35	
0.05PPM	A	Phenolics-	P	0.0524	mg/l	8/20/2025 13:37:36	
0.1PPM	A	Phenolics-	P	0.1032	mg/l	8/20/2025 13:37:37	
0.25PPM	A	Phenolics-	P	0.2535	mg/l	8/20/2025 13:37:38	
0.50PPM	A	Phenolics-	P	0.5003	mg/l	8/20/2025 13:37:39	
1.0PPM	A	Phenolics-	P	0.9835	mg/l	8/20/2025 13:37:40	
2.0PPM	A	Phenolics-	P	2.0075	mg/l	8/20/2025 13:37:41	
ICV1	S	Phenolics-	P	1.0149	mg/l	8/20/2025 14:49:07	
ICB1	S	Phenolics-	P	0.0013	mg/l	8/20/2025 14:49:09	
CCV1	S	Phenolics-	P	0.9884	mg/l	8/20/2025 14:49:10	
CCB1	S	Phenolics-	P	0.0013	mg/l	8/20/2025 14:49:12	
PB169316BL	S	Phenolics-	P	0.0017	mg/l	8/20/2025 14:49:15	
PB169316BS	S	Phenolics-	P	1.0122	mg/l	8/20/2025 14:49:16	
Q2893-10	S	Phenolics-	P	0.027	mg/l	8/20/2025 14:55:53	
Q2893-11	S	Phenolics-	P	0.0531	mg/l	8/20/2025 14:55:55	
CCV2	S	Phenolics-	P	0.9945	mg/l	8/20/2025 14:55:56	
CCB2	S	Phenolics-	P	0.0016	mg/l	8/20/2025 14:55:59	

Alliance Technical Group

284 Sheffield Street, Mountainside, NJ 07092

Reviewed by : RM

Instrument ID : Konelab

8/20/2025 13:44

Test Phenolics-


Accepted 8/20/2025 13:44

Factor 10.07

Bias 0.008

Coeff. of det. 0.999887

Errors



	Calibrator	Response	Calc. con.	Conc.	Errors
1	0.00PPM	0.008	-0.0004	0.0000	-
2	Phenol-2	0.013	0.0524	0.0500	4.8
3	Phenol-2	0.018	0.1032	0.1000	3.2
4	Phenol-2	0.033	0.2535	0.2500	1.4
5	Phenol-2	0.058	0.5003	0.5000	0.1
6	Phenol-2	0.106	0.9835	1.0000	-1.7
7	Phenol-2	0.208	2.0075	2.0000	0.4

08/20/2025
RM