

Analytical Summary Report

Analysis Method: 9045D

Analyst By : jignesh

Parameter: Corrosivity

Supervisor Review By : rubina

Run Number: LB136899

Slope : 99.2

BalanceID: WC SC-7

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W3178
BUFFER PH 7.00 GREEN 1PINT PK6	W3093
PH 10.01 BUFFER,COLOR CD 475ML	W3191
buffer solution pH 7 yellow	W3217
Buffer Solution, PH2 (500ml)	W3161
pH 12.00 Buffer	W3200

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.05].

True Value of CCV2 = 12.00 Control Limits[+/- 0.05].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.2	4.01	08/20/2025	14:15
2	CAL2	1	Water	NA	NA	20.2	7.00	08/20/2025	14:16
3	CAL3	1	Water	NA	NA	20.3	10.02	08/20/2025	14:19
4	ICV	1	Water	NA	NA	20.3	7.02	08/20/2025	14:20
5	CCV1	1	Water	NA	NA	20.2	2.02	08/20/2025	14:25
6	Q2909-02	1	Solid	20.01	20	21.4	5.23	08/20/2025	14:35
7	Q2909-02DUP	1	Solid	20.03	20	21.6	5.25	08/20/2025	14:40
8	CCV2	1	Water	NA	NA	20.3	12.02	08/20/2025	14:44

WORKLIST(Hardcopy Internal Chain)

VB 136899

WorkList Name : corrosivity q2909 WorkList ID : 191376 Department : Wet-Chemistry Date : 08-20-2025 13:49:00

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
Q2909-02	4065	Solid	Corrosivity	Cool 4 deg C	PSEG03	D11	08/20/2025	9045D

Date/Time 08/20/25 13:10 sf
 Raw Sample Received by: sf wyc
 Raw Sample Relinquished by: sf SH

Date/Time 08/20/25 17:30
 Raw Sample Received by: sf SH
 Raw Sample Relinquished by: sf wyc