

Analytical Summary Report

Analysis Method: 1010B

Reviewed By: rubina

Parameter: Flash Point

Supervisor Review By: jignesh

Run Number: LB136907

Ambient Barometric Pressure (mmHg): 755.00

Thermometer ID: Flash Point

Barometric Scale ID: 0511064

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W3194

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	8	28.00	82.4	82.7	08/21/2025	11:10
2	Q2906-01		1	15	100.00	>212.0	>212.0	08/21/2025	11:40
3	Q2906-01DUP		1	15	100.00	>212.0	>212.0	08/21/2025	12:10
4	Q2909-03		1	13	100.00	>212.0	>212.0	08/21/2025	12:40
5	Q2909-04		1	15	100.00	>212.0	>212.0	08/21/2025	13:10

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06

WORKLIST(Hardcopy Internal Chain)

6136907

WorkList Name : fp-8-21

WorkList ID : 191383

Department : Wet-Chemistry

Date : 08-21-2025 08:05:22

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
Q2906-01	3409	Water	Flash Point	Cool 4 deg C	PSEG03	D21	08/20/2025	1010B
Q2909-03	4066	Water	Flash Point	Cool 4 deg C	PSEG03	D11	08/20/2025	1010B
Q2909-04	4067	Water	Flash Point	Cool 4 deg C	PSEG03	D11	08/20/2025	1010B

Date/Time 08/21/2025 11:00
 Raw Sample Received by: RM wcy
 Raw Sample Relinquished by: RM wcy

Date/Time 08/21/2025 13:50
 Raw Sample Received by: RM wcy
 Raw Sample Relinquished by: RM wcy