

LB136911


7470A

INSTRUMENT ID: CV1

Linear

 μ Abs.:

41573



A= 0.0000e+000

B= 2.3923e-004

C= -2.5658e-002

Rho= 0.9999042

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%D
0.0	0.000	-0.041	-0.041	-63	0.000	-63	0				
0.2	0.200	0.184	-0.016	878	0.0 %	878					
2.5	2.500	2.530	0.030	10684	0.0 %	10684	0				
5.0	5.000	5.033	0.033	21147	0.0 %	21147					
7.5	7.500	7.573	0.073	31762	0.0 %	31762					
10.0	10.000	9.920	-0.080	41573	0.0 %	41573					

LB136911 INSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 21 Aug 2025 13:32:44

Type	Sample ID	Extended ID	Conc.	μ Abs.	Units	Type	Std Conc	Method	Date
S	0.0 - 1	50	-	-63	PPB	Std	0.0000	7470A	21 Aug 2025 13:36:4
S	0.2 - 1	50.2	-	878	PPB	Std	0.2000	7470A	21 Aug 2025 13:39:0
S	2.5 - 1	52.5	-	10684	PPB	Std	2.5000	7470A	21 Aug 2025 13:41:2
S	5.0 - 1	55	-	21147	PPB	Std	5.0000	7470A	21 Aug 2025 13:43:4
S	7.5 - 1	57.5	-	31762	PPB	Std	7.5000	7470A	21 Aug 2025 13:45:5
S	10.0 - 1	51.0	-	41573	PPB	Std	10.0000	7470A	21 Aug 2025 13:51:0
U	ICV28 - 1	ICV28	4.0341	16970	PPB	SMPL	-7470A		21 Aug 2025 13:54:0
U	ICB28 - 1	ICB28	-0.0816	-234	PPB	SMPL	-7470A		21 Aug 2025 13:56:2
U	CCV87 - 1	CCV87	5.0965	21411	PPB	SMPL	-7470A		21 Aug 2025 13:58:3
U	CCB87 - 1	CCB87	-0.0757	-209	PPB	SMPL	-7470A		21 Aug 2025 14:00:5
U	CRA - 1	CRA	0.1806	862	PPB	SMPL	-7470A		21 Aug 2025 14:03:0
U	HighStd - 1	HighStd	9.7149	40716	PPB	SMPL	-7470A		21 Aug 2025 14:05:24
U	ChkStd - 1	ChkStd	6.6597	27945	PPB	SMPL	-7470A		21 Aug 2025 14:07:40
U	PB169349BL - 1	PBW	-0.0419	-68	PPB	SMPL	-7470A		21 Aug 2025 14:12:49
U	PB169349BS - 1	LCSW	4.4401	18667	PPB	SMPL	-7470A		21 Aug 2025 14:15:05
U	Q2909-02 - 1	4065	-0.0845	-246	PPB	SMPL	-7470A		21 Aug 2025 14:17:21
U	Q2909-02DUP - 1	4065DUP	-0.0414	-66	PPB	SMPL	-7470A		21 Aug 2025 14:19:39
U	Q2909-02MS - 1	4065MS	4.2207	17750	PPB	SMPL	-7470A		21 Aug 2025 14:21:56
U	Q2909-02MSD - 1	4065MSD	3.7178	15648	PPB	SMPL	-7470A		21 Aug 2025 14:24:13
U	CCV88 - 1	CCV88	5.0219	21099	PPB	SMPL	-7470A		21 Aug 2025 14:26:29
U	CCB88 - 1	CCB88	-0.0831	-240	PPB	SMPL	-7470A		21 Aug 2025 14:28:46
U	PB169325TB - 1	PB169325TB	-0.0465	-87	PPB	SMPL	-7470A		21 Aug 2025 14:31:02
U	Q2909-02,LX5 - 1		-0.0426	-71	PPB	SMPL	-7470A		21 Aug 2025 14:33:17
U	Q2909-02A - 1		4.1621	17505	PPB	SMPL	-7470A		21 Aug 2025 14:38:06
U	CCV89 - 1	CCV89	5.0238	21107	PPB	SMPL	-7470A		21 Aug 2025 14:40:22
U	CCB89 - 1	CCB89	-0.0855	-250	PPB	SMPL	-7470A		21 Aug 2025 14:42:39