

LB136912

=====

Test results

Aquakem 7.2AQ1

Page: 1

Alliance Technical Group

284 Sheffield Street, Mountainside, NJ 07092

Reviewed by : RM Instrument ID : Konelab

8/21/2025 13:29

Test: Phenolics-

Sample Id	Result	Dil. 1 +	Response	Errors
ICV1	0.999	0.0	0.108	
ICB1	0.001	0.0	0.008	
CCV1	0.978	0.0	0.106	
CCB1	0.001	0.0	0.008	
PB169343BL	0.001	0.0	0.008	
PB169343BS	1.006	0.0	0.109	
Q2893-03	0.025	0.0	0.011	
PB169345BL	0.002	0.0	0.008	
PB169345BS	1.009	0.0	0.109	
Q2893-09	0.025	0.0	0.011	
CCV2	0.978	0.0	0.106	
CCB2	0.001	0.0	0.008	

N	12
Mean	0.419
SD	0.5077
CV%	121.25

Aquakem v. 7.2AQ1

Results from time period:

Thu Aug 21 12:44:22 2025

Thu Aug 21 13:29:07 2025

Sample Id	Sam/Ctr/c	Test short r	Test type	Result	Result unit	Result date and time	Stat
0.00PPM	A	Phenolics-	P	0.0009	mg/l	8/21/2025 12:44:22	
0.05PPM	A	Phenolics-	P	0.0531	mg/l	8/21/2025 12:44:23	
0.1PPM	A	Phenolics-	P	0.1039	mg/l	8/21/2025 12:44:24	
0.25PPM	A	Phenolics-	P	0.2502	mg/l	8/21/2025 12:44:25	
0.50PPM	A	Phenolics-	P	0.4988	mg/l	8/21/2025 12:44:26	
1.0PPM	A	Phenolics-	P	0.9862	mg/l	8/21/2025 12:44:27	
2.0PPM	A	Phenolics-	P	2.0069	mg/l	8/21/2025 12:44:28	
ICV1	S	Phenolics-	P	0.9988	mg/l	8/21/2025 13:14:08	
ICB1	S	Phenolics-	P	0.0014	mg/l	8/21/2025 13:14:11	
CCV1	S	Phenolics-	P	0.9784	mg/l	8/21/2025 13:14:12	
CCB1	S	Phenolics-	P	0.0012	mg/l	8/21/2025 13:14:15	
PB169343BL	S	Phenolics-	P	0.0009	mg/l	8/21/2025 13:14:16	
PB169343BS	S	Phenolics-	P	1.0056	mg/l	8/21/2025 13:14:18	
Q2893-03	S	Phenolics-	P	0.0251	mg/l	8/21/2025 13:24:10	
PB169345BL	S	Phenolics-	P	0.0018	mg/l	8/21/2025 13:24:15	
PB169345BS	S	Phenolics-	P	1.0085	mg/l	8/21/2025 13:24:17	
Q2893-09	S	Phenolics-	P	0.0246	mg/l	8/21/2025 13:24:18	
CCV2	S	Phenolics-	P	0.9775	mg/l	8/21/2025 13:29:05	
CCB2	S	Phenolics-	P	0.0012	mg/l	8/21/2025 13:29:06	

Alliance Technical Group

284 Sheffield Street, Mountainside, NJ 07092

Reviewed by : RM

Instrument ID : Konelab

8/21/2025 12:50


Test Phenolics-

Accepted 8/21/2025 12:50

Factor 9.99
Bias 0.008

Coeff. of det. 0.999915

Errors



	Calibrator	Response	Calc. con.	Conc.	Errors
1	0.00PPM	0.008	0.0009	0.0000	6.2
2	Phenol-2	0.014	0.0531	0.0500	3.9
3	Phenol-2	0.019	0.1039	0.1000	0.1
4	Phenol-2	0.033	0.2502	0.2500	-0.2
5	Phenol-2	0.058	0.4988	0.5000	-1.4
6	Phenol-2	0.107	0.9862	1.0000	0.3
7	Phenol-2	0.209	2.0069	2.0000	

08/21/2025
RM