

16136913

Test results

Aquakem 7.2AQ1

Page: 1

Alliance Technical Group

284 Sheffield Street, Mountainside, NJ 07092

Reviewed by : RM

Instrument ID : Konelab

8/21/2025 14:02

Test: Phenolics-

Sample Id	Result	Dil. 1 +	Response	Errors
ICV1	1.008	0.0	0.109	
ICB1	0.003	0.0	0.009	
CCV1	0.977	0.0	0.106	
CCB1	0.002	0.0	0.008	
PB169344BL	0.002	0.0	0.008	
PB169344BS	1.009	0.0	0.109	
Q2893-12	0.026	0.0	0.011	
CCV2	0.979	0.0	0.106	
CCB2	0.003	0.0	0.009	

N 9
Mean 0.445
SD 0.5199
CV% 116.73

Aquakem v. 7.2AQ1

Results from time period:

Thu Aug 21 13:54:51 2025

Thu Aug 21 14:01:44 2025

Sample Id	Sam/Ctr/c/	Test short r	Test type	Result	Result unit	Result date and time	Stat
0.00PPM	A	Phenolics-	P	0.0009	mg/l	8/21/2025 12:44:22	
0.05PPM	A	Phenolics-	P	0.0531	mg/l	8/21/2025 12:44:23	
0.1PPM	A	Phenolics-	P	0.1039	mg/l	8/21/2025 12:44:24	
0.25PPM	A	Phenolics-	P	0.2502	mg/l	8/21/2025 12:44:25	
0.50PPM	A	Phenolics-	P	0.4988	mg/l	8/21/2025 12:44:26	
1.0PPM	A	Phenolics-	P	0.9862	mg/l	8/21/2025 12:44:27	
2.0PPM	A	Phenolics-	P	2.0069	mg/l	8/21/2025 12:44:28	
ICV1	S	Phenolics-	P	1.0075	mg/l	8/21/2025 13:54:51	
ICB1	S	Phenolics-	P	0.0031	mg/l	8/21/2025 13:54:53	
CCV1	S	Phenolics-	P	0.9774	mg/l	8/21/2025 13:54:55	
CCB1	S	Phenolics-	P	0.0015	mg/l	8/21/2025 13:54:58	
PB169344BL	S	Phenolics-	P	0.0024	mg/l	8/21/2025 13:55:00	
PB169344BS	S	Phenolics-	P	1.0092	mg/l	8/21/2025 13:55:02	
Q2893-12	S	Phenolics-	P	0.0257	mg/l	8/21/2025 14:01:37	
CCV2	S	Phenolics-	P	0.9788	mg/l	8/21/2025 14:01:41	
CCB2	S	Phenolics-	P	0.0029	mg/l	8/21/2025 14:01:44	

Alliance Technical Group
 284 Sheffield Street, Mountainside, NJ 07092

8/21/2025 12:50 Reviewed by : RM Instrument ID : Konelab


Test Phenolics-

Accepted 8/21/2025 12:50

Factor 9.99
 Bias 0.008

Coeff. of det. 0.999915

Errors



	Calibrator	Response	Calc. con.	Conc.	Errors
1	0.00PPM	0.008	0.0009	0.0000	-
2	Phenol-2	0.014	0.0531	0.0500	6.2
3	Phenol-2	0.019	0.1039	0.1000	3.9
4	Phenol-2	0.033	0.2502	0.2500	0.1
5	Phenol-2	0.058	0.4988	0.5000	-0.2
6	Phenol-2	0.107	0.9862	1.0000	-1.4
7	Phenol-2	0.209	2.0069	2.0000	0.3

08/21/2025
 RM