

LB136922


7471B

INSTRUMENT ID: CV1

Linear

 μ Abs.:

52496



A= 0.0000e+000

B= 1.9354e-004

C= 1.6798e-002

Rho= 0.9990738

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%D
0.0	0.000	0.017	0.017	0	0.000	0					1
0.2	0.200	0.204	0.004	967	0.0 %	967					2
2.5	2.500	2.542	0.042	13049	0.0 %	13049					2
5.0	5.000	5.091	0.091	26217	0.0 %	26217					2
7.5	7.500	7.169	-0.331	36954	0.0 %	36954					-4
10.0	10.000	10.177	0.177	52496	0.0 %	52496					2

LB136922 INSTRUMENT ID : CV1

Method: 7471B		Operator: Admin		Date of Analysis: 22 Aug 2025 11:47:34					
Type	Sample ID	Extended ID	Conc.	µ Abs.	Units	Type	Std Conc	Method	Date
S	0.0 - 1	50.0	-	0	PPB	Std	0.0000	7471B	22 Aug 2025 12:22:49
S	0.2 - 1	50.2	-	967	PPB	Std	0.2000	7471B	22 Aug 2025 12:25:15
S	2.5 - 1	52.5	-	13049	PPB	Std	2.5000	7471B	22 Aug 2025 12:27:31
S	5.0 - 1	55.0	-	26217	PPB	Std	5.0000	7471B	22 Aug 2025 12:29:48
S	7.5 - 1	57.5	-	36954	PPB	Std	7.5000	7471B	22 Aug 2025 12:32:05
S	10.0 - 1	510.0	-	52496	PPB	Std	10.0000	7471B	22 Aug 2025 12:37:26
U	ICV01 - 1	ICV01	3.9898	20528	PPB	SMPL	-	7471B	22 Aug 2025 12:41:32
U	ICB01 - 1	ICB01	-0.0244	-213	PPB	SMPL	-	7471B	22 Aug 2025 12:43:47
U	CCV01 - 1	CCV01	4.9864	25677	PPB	SMPL	-	7471B	22 Aug 2025 12:46:04
U	CCB01 - 1	CCB01	-0.0281	-232	PPB	SMPL	-	7471B	22 Aug 2025 12:48:19
U	CRA - 1	CRA	0.2109	1003	PPB	SMPL	-	7471B	22 Aug 2025 12:50:36
U	HighStd - 1	HighStd	10.0518	51849	PPB	SMPL	-	7471B	22 Aug 2025 12:52:51
U	ChkStd - 1	ChkStd	6.7403	34739	PPB	SMPL	-	7471B	22 Aug 2025 12:55:06
U	PB169351BL - 1	PB169351BL	-0.0252	-217	PPB	SMPL	-	7471B	22 Aug 2025 12:57:25
U	PB169351BS - 1	PB169351BS	4.4108	22703	PPB	SMPL	-	7471B	22 Aug 2025 13:00:13
U	Q2909-01 - 1	4065	0.3166	1549	PPB	SMPL	-	7471B	22 Aug 2025 13:02:29
U	Q2915-01 - 1	BRIDGE-WEST	0.2659	1287	PPB	SMPL	-	7471B	22 Aug 2025 13:04:53
U	Q2915-03 - 1	BRIDGE-EAST	0.4021	1991	PPB	SMPL	-	7471B	22 Aug 2025 13:07:18
U	Q2916-01 - 1	EO-03-082025	0.7656	3869	PPB	SMPL	-	7471B	22 Aug 2025 13:09:37
U	Q2917-01 - 1	OR-02-082025	5.5032	28347	PPB	SMPL	-	7471B	22 Aug 2025 13:11:55
U	CCV02 - 1	CCV02	4.9862	25676	PPB	SMPL	-	7471B	22 Aug 2025 13:17:08
U	CCB02 - 1	CCB02	-0.0173	-176	PPB	SMPL	-	7471B	22 Aug 2025 13:19:26
U	Q2917-01DUP - 1	OR-02-082025DUP	5.4515	28080	PPB	SMPL	-	7471B	22 Aug 2025 13:24:59
U	Q2917-01MS - 1	OR-02-082025MS	9.0684	46768	PPB	SMPL	-	7471B	22 Aug 2025 13:30:47
U	Q2917-01MSD - 1	OR-02-082025MSD	8.8184	45476	PPB	SMPL	-	7471B	22 Aug 2025 13:33:02
U	Q2919-01 - 1	OK-02-082025	0.5558	2785	PPB	SMPL	-	7471B	22 Aug 2025 13:35:21
U	Q2932-01 - 1	ARS20-0009	0.3241	1588	PPB	SMPL	-	7471B	22 Aug 2025 13:37:40
U	Q2932-03 - 1	ARS20-0002-SOIL	0.5158	2578	PPB	SMPL	-	7471B	22 Aug 2025 13:39:58
U	Q2933-02 - 1	25-SEPTIC	0.1542	710	PPB	SMPL	-	7471B	22 Aug 2025 13:42:14
U	Q2933-04 - 1	29-SEPTIC	0.1993	943	PPB	SMPL	-	7471B	22 Aug 2025 13:44:31
U	Q2917-01LX5 - 1		1.0544	5361	PPB	SMPL	-	7471B	22 Aug 2025 13:46:48
U	Q2917-01A - 1		9.5509	49261	PPB	SMPL	-	7471B	22 Aug 2025 13:49:03
U	CCV03 - 1	CCV03	5.0272	25888	PPB	SMPL	-	7471B	22 Aug 2025 13:51:19
U	CCB03 - 1	CCB03	0.0141	-14	PPB	SMPL	-	7471B	22 Aug 2025 13:56:43