

Analytical Summary Report

Analysis Method: 1010B

Reviewed By: Eman

Parameter: Flash Point

Supervisor Review By: jignesh

Run Number: LB136927

Ambient Barometric Pressure (mmHg): 760.00

Thermometer ID: Flashpoint

Barometric Scale ID: 0511064

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W3194

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	9	28.00	82.4	82.4	08/22/2025	09:00
2	Q2927-01		1	14	100.00	>212.0	>212.0	08/22/2025	09:30
3	Q2927-01DUP		1	14	100.00	>212.0	>212.0	08/22/2025	10:00

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06

WORKLIST(Hardcopy Internal Chain)

16136927

WorkList Name : fp-08-22

WorkList ID : 191419

Department : Wet-Chemistry

Date : 08-22-2025 08:14:57

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
Q2927-01	A6121M	Water	Flash Point	Cool 4 deg C	PSEG03	D31	08/21/2025	1010B

Date/Time 08/22/25 08:50
 Raw Sample Received by: EM(WC)
 Raw Sample Relinquished by: EM(WC)

Date/Time 08/22/25 10:40
 Raw Sample Received by: EM(WC)
 Raw Sample Relinquished by: EM(WC)