

LB13B932


7470A

INSTRUMENT ID: CV1

Linear

 μ Abs.:

38877



A= 0.0000e+000

B= 2.5115e-004

C= -4.4602e-002

Rho= 0.9981448

Accept= Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%D
0.0	0.000	-0.044	-0.044	3	0.000	3	0				-7
0.2	0.200	0.187	-0.013	923	0.0 %	923					4
2.5	2.500	2.596	0.096	10514	0.0 %	10514	0				4
5.0	5.000	4.817	-0.183	19359	0.0 %	19359					8
7.5	7.500	7.924	0.424	31727	0.0 %	31727					3
10.0	10.000	9.719	-0.281	38877	0.0 %	38877					

LB136932 INSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 22 Aug 2025 14:44:34

Type	Sample ID	Extended ID	Conc.	μ Abs.	Units	Type	Std ConcMethod	Date
S	0.0 - 1	50	-	3	PPB	Std	0.00007470A	22 Aug 2025 15:26:50
S	0.2 - 1	50.2	-	923	PPB	Std	0.20007470A	22 Aug 2025 15:29:14
S	2.5 - 1	52.5	-	10514	PPB	Std	2.50007470A	22 Aug 2025 15:31:30
S	5.0 - 1	55	-	19359	PPB	Std	5.00007470A	22 Aug 2025 15:33:47
S	7.5 - 1	57.5	-	31727	PPB	Std	7.50007470A	22 Aug 2025 15:36:15
S	10.0 - 1	511	-	38877	PPB	Std	10.00007470A	22 Aug 2025 15:38:31
U	ICV02 - 1	ICV02	4.1567	16728	PPB	SMPL	-7470A	22 Aug 2025 15:43:06
U	ICB02 - 1	ICB02	-0.0828	-152	PPB	SMPL	-7470A	22 Aug 2025 15:45:20
U	CCV04 - 1	CCV04	4.9619	19934	PPB	SMPL	-7470A	22 Aug 2025 15:47:36
U	CCB04 - 1	CCB04	-0.0911	-185	PPB	SMPL	-7470A	22 Aug 2025 15:49:51
U	CRA - 1	CRA	0.2058	997	PPB	SMPL	-7470A	22 Aug 2025 15:54:48
U	HighStd - 1	HighStd	9.6481	38593	PPB	SMPL	-7470A	22 Aug 2025 15:57:03
U	ChkStd - 1	ChkStd	7.0108	28092	PPB	SMPL	-7470A	22 Aug 2025 15:59:18
U	PB169363BL - 1	PBW	-0.0936	-195	PPB	SMPL	-7470A	22 Aug 2025 16:01:36
U	PB169363BS - 1	LCSW	4.6602	18733	PPB	SMPL	-7470A	22 Aug 2025 16:03:53
U	Q2927-01 - 1	A6121M	-0.0572	-50	PPB	SMPL	-7470A	22 Aug 2025 16:06:09
U	Q2927-01DUP - 1	A6121MDUP	0.0036	192	PPB	SMPL	-7470A	22 Aug 2025 16:08:27
U	Q2927-01MS - 1	A6121MMS	3.4185	13789	PPB	SMPL	-7470A	22 Aug 2025 16:10:44
U	Q2927-01MSD - 1	A6121MMSD	3.3023	13326	PPB	SMPL	-7470A	22 Aug 2025 16:13:08
U	CCV05 - 1	CCV05	5.4034	21692	PPB	SMPL	-7470A	22 Aug 2025 16:15:26
U	CCB05 - 1	CCB05	-0.0499	-21	PPB	SMPL	-7470A	22 Aug 2025 16:20:22
U	Q2927-01LX5 - 1		-0.0403	17	PPB	SMPL	-7470A	22 Aug 2025 16:22:37
U	Q2927-01A - 1		4.3059	17322	PPB	SMPL	-7470A	22 Aug 2025 16:24:52
U	CCV06 - 1	CCV06	4.8820	19616	PPB	SMPL	-7470A	22 Aug 2025 16:29:45
U	CCB06 - 1	CCB06	-0.0973	-210	PPB	SMPL	-7470A	22 Aug 2025 16:32:00