

Analytical Summary Report

Analysis Method: 1010B

Reviewed By: rubina

Parameter: Flash Point

Supervisor Review By: jignesh

Run Number: LB136948

Ambient Barometric Pressure (mmHg): 755.00

Thermometer ID: Flash Point

Barometric Scale ID: 0511064

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W3194

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	9	28.00	82.4	82.7	08/25/2025	10:20
2	Q2940-01		1	14	100.00	>212.0	>212.0	08/25/2025	11:30
3	Q2940-01DUP		1	14	100.00	>212.0	>212.0	08/25/2025	12:00
4	Q2944-01		1	12	100.00	>212.0	>212.0	08/25/2025	12:30
5	Q2944-02		1	15	100.00	>212.0	>212.0	08/25/2025	13:00

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06

WORKLIST(Hardcopy Internal Chain)

6136948

WorkList Name : fp-8-25

WorkList ID : 191443

Department : Wet-Chemistry

Date : 08-25-2025 08:37:04

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
Q2940-01	COMP(1-6)	Water	Flash Point	Cool 4 deg C	PSEG03	D21	08/22/2025	1010B
Q2944-01	AUD-25-0145	Water	Flash Point	Cool 4 deg C	PSEG03	D31	08/22/2025	1010B
Q2944-02	AUD-25-0146	Water	Flash Point	Cool 4 deg C	PSEG03	D31	08/22/2025	1010B

Date/Time 08/25/2025 10:10
Raw Sample Received by: RM (wsc)
Raw Sample Relinquished by: RM (wsc)

Date/Time 08/25/2025 13:10
Raw Sample Received by: RM (wsc)
Raw Sample Relinquished by: RM (wsc)