

Analytical Summary Report

Analysis Method: 9040C

Parameter: pH

Run Number: LB136956

Analyst By : jignesh

Supervisor Review By : rubina

Slope : 98.6

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W3178
BUFFER PH 7.00 GREEN 1PINT PK6	W3093
PH 10.01 BUFFER,COLOR CD 475ML	W3191
buffer solution pH 7 yellow	W3217
Buffer Solution, PH2 (500ml)	W3161
pH 12.00 Buffer	W3200

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.05].

True Value of CCV2 = 12.00 Control Limits[+/- 0.05].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.2	4.01	08/26/2025	10:15
2	CAL2	1	Water	NA	NA	20.2	7.00	08/26/2025	10:16
3	CAL3	1	Water	NA	NA	20.3	10.02	08/26/2025	10:19
4	ICV	1	Water	NA	NA	20.3	7.00	08/26/2025	10:22
5	CCV1	1	Water	NA	NA	20.2	2.02	08/26/2025	10:25
6	Q2954-11	1	Water	NA	NA	22.6	6.17	08/26/2025	10:35
7	Q2955-02	1	Water	NA	NA	20.2	4.87	08/26/2025	10:45
8	Q2955-02DUP	1	Water	NA	NA	20.5	4.89	08/26/2025	10:55
9	CCV2	1	Water	NA	NA	20.3	12.02	08/26/2025	11:00

WORKLIST(Hardcopy Internal Chain)

VB 126956

WorkList Name : ph wq2654

WorkList ID : 191483

Department : Wet-Chemistry

Date : 08-26-2025 09:54:53

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
Q2954-11	BRIDGE	Water	pH	Cool 4 deg C	PSEG03	D41	08/25/2025	9040C
Q2955-02	OILY-WATER	Water	pH	Cool 4 deg C	PSEG03	D41	08/25/2025	9040C

Date/Time 08/26/25 10:10
 Raw Sample Received by: JB wlc
 Raw Sample Relinquished by: JB wlc

Date/Time 08/26/25 13:00
 Raw Sample Received by: JB wlc
 Raw Sample Relinquished by: JB wlc