

Analytical Summary Report

Analysis Method: 9045D

Analyst By : jignesh

Parameter: Corrosivity

Supervisor Review By : rubina

Run Number: LB136969

Slope : 99.2

BalanceID: WC SC-7

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W3178
BUFFER PH 7.00 GREEN 1PINT PK6	W3093
PH 10.01 BUFFER,COLOR CD 475ML	W3191
buffer solution pH 7 yellow	W3217
Buffer Solution, PH2 (500ml)	W3161
pH 12.00 Buffer	W3200

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.05].

True Value of CCV2 = 12.00 Control Limits[+/- 0.05].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.2	4.01	08/26/2025	15:00
2	CAL2	1	Water	NA	NA	20.2	7.00	08/26/2025	15:01
3	CAL3	1	Water	NA	NA	20.2	10.03	08/26/2025	15:05
4	ICV	1	Water	NA	NA	20.3	7.02	08/26/2025	15:10
5	CCV1	1	Water	NA	NA	20.2	2.01	08/26/2025	15:15
6	Q2955-01	1	Solid	20.02	20	22.4	6.23	08/26/2025	15:30
7	Q2955-01DUP	1	Solid	20.04	20	22.6	6.25	08/26/2025	15:33
8	CCV2	1	Water	NA	NA	20.3	12.02	08/26/2025	15:35

WORKLIST(Hardcopy Internal Chain)

136969

WorkList Name : corrosivity q2959 WorkList ID : 191504 Department : Wet-Chemistry Date : 08-26-2025 14:25:00

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
Q2955-01	OILY-GRAVEL	Solid	Corrosivity	Cool 4 deg C	PSEG03	D41	08/25/2025	9045D

Date/Time 08/26/25 14:35
Raw Sample Received by: JDCSM
Raw Sample Relinquished by: JDCSM

Date/Time 08/26/25 17:30
Raw Sample Received by: JDCSM
Raw Sample Relinquished by: JDCSM