

LB136975


7470A

INSTRUMENT ID: CV1

Linear

 μ Abs.:

41458



A= 0.0000e+000

B= 2.3723e-004

C= 3.8862e-002

Rho= 0.9995391

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	0/0.0
0.0	0.000	0.037	0.037	-8	0.000	-8	0				-
0.2	0.200	0.219	0.019	759	0.0 %	759					10
2.5	2.500	2.369	-0.131	9823	0.0 %	9823	0				-5
5.0	5.000	5.000	-0.000	20911	0.0 %	20911					0
7.5	7.500	7.701	0.201	32300	0.0 %	32300					2
10.0	10.000	9.874	-0.126	41458	0.0 %	41458					-1

LB136975 INSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 27 Aug 2025 09:03:07

Type	Sample ID	Extended ID	Conc.	μ Abs.	Units	Type	Std ConcMethod	Date
S	0.0 - 1	50	-	-8	PPB	Std	0.00007470A	27 Aug 2025 09:43:09
S	0.2 - 1	50.2	-	759	PPB	Std	0.20007470A	27 Aug 2025 09:45:34
S	2.5 - 1	52.5	-	9823	PPB	Std	2.50007470A	27 Aug 2025 09:47:51
S	5.0 - 1	55	-	20911	PPB	Std	5.00007470A	27 Aug 2025 09:52:51
S	7.5 - 1	57.5	-	32300	PPB	Std	7.50007470A	27 Aug 2025 09:57:44
S	10.0 - 1	510	-	41458	PPB	Std	10.00007470A	27 Aug 2025 10:00:00
U	ICV07 - 1	ICV07	4.2144	17601	PPB	SMPL	-7470A	27 Aug 2025 10:03:11
U	ICB07 - 1	ICB07	0.0578	80	PPB	SMPL	-7470A	27 Aug 2025 10:05:25
U	CCV18 - 1	CCV18	5.0183	20990	PPB	SMPL	-7470A	27 Aug 2025 10:07:42
U	CCB18 - 1	CCB18	0.0737	147	PPB	SMPL	-7470A	27 Aug 2025 10:09:57
U	CRA - 1	CRA	0.2272	794	PPB	SMPL	-7470A	27 Aug 2025 10:12:13
U	HighStd - 1	HighStd	9.8804	41485	PPB	SMPL	-7470A	27 Aug 2025 10:14:38
U	ChkStd - 1	ChkStd	7.0818	29688	PPB	SMPL	-7470A	27 Aug 2025 10:17:10
U	PB169413BL - 1	PBW	-0.0389	-328	PPB	SMPL	-7470A	27 Aug 2025 10:19:28
U	PB169413BS - 1	LCSW	4.3361	18114	PPB	SMPL	-7470A	27 Aug 2025 10:21:46
U	Q2934-01 - 1	ENV-SW01	0.0272	-49	PPB	SMPL	-7470A	27 Aug 2025 10:24:02
U	Q2934-02 - 1	ENV-SW02	0.0906	218	PPB	SMPL	-7470A	27 Aug 2025 10:26:20
U	Q2934-03 - 1	ENV-SW03	0.1129	312	PPB	SMPL	-7470A	27 Aug 2025 10:28:37
U	Q2934-04 - 1	ENV-SW04	0.1409	430	PPB	SMPL	-7470A	27 Aug 2025 10:30:54
U	Q2940-01 - 1	COMP(1-6)	0.0296	-39	PPB	SMPL	-7470A	27 Aug 2025 10:33:09
U	CCV19 - 1	CCV19	4.9056	20515	PPB	SMPL	-7470A	27 Aug 2025 10:35:24
U	CCB19 - 1	CCB19	0.0386	-1	PPB	SMPL	-7470A	27 Aug 2025 10:37:39
U	Q2940-01DUP - 1	COMP(1-6)DUP	0.0488	42	PPB	SMPL	-7470A	27 Aug 2025 10:39:55
U	Q2940-01MS - 1	COMP(1-6)MS	4.1918	17506	PPB	SMPL	-7470A	27 Aug 2025 10:42:11
U	Q2940-01MSD - 1	COMP(1-6)MSD	4.2072	17571	PPB	SMPL	-7470A	27 Aug 2025 10:44:26
U	Q2954-11 - 1	BRIDGE	0.8699	3503	PPB	SMPL	-7470A	27 Aug 2025 10:46:43
U	Q2961-01 - 1	4075	0.4671	1805	PPB	SMPL	-7470A	27 Aug 2025 10:49:00
U	Q2940-01LX5 - 1		0.0538	63	PPB	SMPL	-7470A	27 Aug 2025 11:26:18
U	Q2940-01A - 1		3.2424	13504	PPB	SMPL	-7470A	27 Aug 2025 11:28:34
U	CCV20 - 1	CCV20	5.3077	22210	PPB	SMPL	-7470A	27 Aug 2025 11:33:32
U	CCB20 - 1	CCB20	0.0389	0	PPB	SMPL	-7470A	27 Aug 2025 11:35:49