

Analytical Summary Report

Analysis Method: 9040C

Parameter: pH

Run Number: LB136977

Analyst By : jignesh

Supervisor Review By : rubina

Slope : 98.6

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W3178
BUFFER PH 7.00 GREEN 1PINT PK6	W3093
PH 10.01 BUFFER,COLOR CD 475ML	W3191
buffer solution pH 7 yellow	W3217
Buffer Solution, PH2 (500ml)	W3161
pH 12.00 Buffer	W3200

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.05].

True Value of CCV2 = 12.00 Control Limits[+/- 0.05].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.3	4.01	08/27/2025	10:30
2	CAL2	1	Water	NA	NA	20.3	7.01	08/27/2025	10:32
3	CAL3	1	Water	NA	NA	20.3	10.02	08/27/2025	10:35
4	ICV	1	Water	NA	NA	20.3	7.00	08/27/2025	10:36
5	CCV1	1	Water	NA	NA	20.3	2.01	08/27/2025	10:38
6	Q2961-01	1	Water	NA	NA	21.9	7.66	08/27/2025	10:45
7	Q2961-01DUP	1	Water	NA	NA	21.6	7.67	08/27/2025	10:47
8	CCV2	1	Water	NA	NA	20.3	12.02	08/27/2025	10:50

VB 136977

WORKLIST(Hardcopy Internal Chain)

WorkList Name : PH W Q2961 WorkList ID : 191518 Department : Wet-Chemistry Date : 08-27-2025 10:04:31

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
Q2961-01	B 4075	Water	pH	Cool 4 deg C	PSEG03	D31	08/26/2025	9040C

Date/Time 08/27/25 10:15
Raw Sample Received by: JB Cole
Raw Sample Relinquished by: JDCSM

Date/Time 08/27/25 12:30
Raw Sample Received by: JDCSM
Raw Sample Relinquished by: JB Cole