

Analytical Summary Report

Analysis Method: 9040C

Parameter: pH

Run Number: LB136990

Analyst By : jignesh

Supervisor Review By : rubina

Slope : 98.2

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W3178
BUFFER PH 7.00 GREEN 1PINT PK6	W3093
PH 10.01 BUFFER,COLOR CD 475ML	W3191
buffer solution pH 7 yellow	W3217
Buffer Solution, PH2 (500ml)	W3161
pH 12.00 Buffer	W3200

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.05].

True Value of CCV2 = 12.00 Control Limits[+/- 0.05].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.2	4.01	08/28/2025	07:45
2	CAL2	1	Water	NA	NA	20.2	7.00	08/28/2025	07:46
3	CAL3	1	Water	NA	NA	20.3	10.02	08/28/2025	07:48
4	ICV	1	Water	NA	NA	20.3	7.00	08/28/2025	08:00
5	CCV1	1	Water	NA	NA	20.2	2.02	08/28/2025	08:01
6	Q2965-01	1	Water	NA	NA	20.9	7.40	08/28/2025	08:15
7	Q2971-37	1	Water	NA	NA	20.2	6.03	08/28/2025	08:25
8	Q2971-37DUP	1	Water	NA	NA	20.5	6.05	08/28/2025	08:33
9	CCV2	1	Water	NA	NA	20.2	12.02	08/28/2025	08:37

13136990

WORKLIST(Hardcopy Internal Chain)

WorkList Name : ph w q2971

WorkList ID : 191538

Department : Wet-Chemistry

Date : 08-28-2025 07:34:48

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
Q2965-01	C 50788	Water	pH	Cool 4 deg C	PSEG03	D31	08/27/2025	9040C
Q2971-37	K GROUNDWATER	Water	pH	Cool 4 deg C	PSEG03	D31	08/27/2025	9040C

Date/Time 08/28/25 07:40

Raw Sample Received by: SP WWC

Raw Sample Relinquished by: JDCSM

Date/Time 08/28/25

Raw Sample Received by: JDCSM

Raw Sample Relinquished by: SP WWC