

Analytical Summary Report

Analysis Method: SM5220 D

ANALYST: Eman

Parameter: COD

SUPERVISOR REVIEW BY: jignesh

Run Number: LB136994

Reagent/Standard	Lot/Log #
COD calibration std. 150 ppm	WP113238
COD calibration std. 100 ppm	WP113237
COD calibration std. 50 ppm	WP113235
COD calibration std. 10 ppm	WP113234
COD calibration std. 0 ppm	WP113233
COD ICV-LCS std, 50ppm	WP113240
COD calibration std. 75 ppm	WP113236
COD CCV std, 50ppm	WP114380
COD ICV-LCS std, 50ppm	WP114381
RL CHECK	WP114382
COD Digestion Vials Low Level 0-150Mg/L	W3229

Temp In(C): 149	Date In: 08/28/2025	Time In: 08:40
Temp Out(C): 152	Date Out: 08/28/2025	Time Out: 10:40

Intercept: 0.8179

Slope: 0.9847

Regression: 0.9995

Seq	Lab ID	TrueValue (mg/l)	DF	MATRIX	Reading	Result (mg/l)	%D	Anal Date	Anal Time
1	CAL1	0	1	Water	0.000	-0.831		05/28/2025	13:10
2	CAL2	10	1	Water	9.000	8.309	-16.9	05/28/2025	13:10
3	CAL3	50	1	Water	52.000	51.977	4	05/28/2025	13:11
4	CAL4	75	1	Water	77.000	77.366	3.2	05/28/2025	13:11
5	CAL5	100	1	Water	99.000	99.708	-0.3	05/28/2025	13:12
6	CAL6	150	1	Water	147.000	148.453	-1	05/28/2025	13:12

Analytical Summary Report

Analysis Method: SM5220 D

ANALYST: Eman

Parameter: COD

SUPERVISOR REVIEW BY: jignesh

Run Number: LB136994

Seq	Lab ID	True Value (mg/l)	Initial Weight (g)	Final Vol (ml)	DF	MATRIX	Reading	Result	AnalDate	AnalTime
1	ICV	50	NA	NA	1	Water	51.000	50.962	05/28/2025	13:13
2	ICB		NA	NA	1	Water	0.000	-0.831	05/28/2025	13:13
3	CCV1	50	NA	NA	1	Water	50.000	49.946	08/28/2025	11:20
4	CCB1		NA	NA	1	Water	1.000	0.185	08/28/2025	11:20
5	RL Check	10	NA	NA	1	Water	10.000	9.325	08/28/2025	11:21
6	LB136994BL		NA	NA	1	Water	0.000	-0.831	08/28/2025	11:21
7	LB136994BS	50	NA	NA	1	Water	49.000	48.931	08/28/2025	11:22
8	Q2971-37		NA	NA	1	Water	11.000	10.340	08/28/2025	11:22
9	Q2971-37DUP		NA	NA	1	Water	12.000	11.356	08/28/2025	11:23
10	Q2971-37MS	50	NA	NA	1	Water	63.000	63.148	08/28/2025	11:23
11	Q2971-37MSD	50	NA	NA	1	Water	61.000	61.117	08/28/2025	11:24
12	CCV2	50	NA	NA	1	Water	51.000	50.962	08/28/2025	11:24
13	CCB2		NA	NA	1	Water	0.000	-0.831	08/28/2025	11:25

1b136994

WORKLIST(Hardcopy Internal Chain)

WorkList Name : COD-082725

WorkList ID : 191554

Department : Wet-Chemistry

Date : 08-27-2025 16:00:35

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
Q2971-37	GROUNDWATER	Water	COD	Conc H2SO4 to pH < 2	PSEG03	D31	08/27/2025	SM5220 D

Date/Time

08/28/25 08:05

Raw Sample Received by:

Em(WC)

Raw Sample Relinquished by:

Em(WC)

Date/Time

08/28/25

Raw Sample Received by:

Em(WC)

Raw Sample Relinquished by:

Em(WC)