

LB137014


7470A

INSTRUMENT ID: CV1

Linear

 μ Abs.:

40976



A= 0.0000e+000

B= 2.4112e-004

C= -1.8917e-002

Rho= 0.9993461

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%D
0.0	0.000	-0.002	-0.002	70	0.000	70	0				-3
0.2	0.200	0.206	0.006	931	0.0 %	931					3
2.5	2.500	2.518	0.018	10521	0.0 %	10521	0				1
5.0	5.000	4.858	-0.142	20226	0.0 %	20226					-3
7.5	7.500	7.759	0.259	32257	0.0 %	32257					3
10.0	10.000	9.861	-0.139	40976	0.0 %	40976					-1

LB137014 INSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 29 Aug 2025 12:21:48

Type	Sample ID	Extended ID	Conc.	μ Abs.	Units	Type	Std ConcMethod	Date
S	0.0 - 1	50	-	70	PPB	Std	0.00007470A	29 Aug 2025 12:27:17
S	0.2 - 1	50.2	-	931	PPB	Std	0.20007470A	29 Aug 2025 12:29:34
S	2.5 - 1	52.5	-	10521	PPB	Std	2.50007470A	29 Aug 2025 12:31:52
S	5.0 - 1	55	-	20226	PPB	Std	5.00007470A	29 Aug 2025 12:34:08
S	7.5 - 1	57.5	-	32257	PPB	Std	7.50007470A	29 Aug 2025 12:36:25
S	10.0 - 1	51	-	40976	PPB	Std	10.00007470A	29 Aug 2025 12:43:11
U	ICV14 - 1	ICV14	3.8974	16242	PPB	SMPL	-7470A	29 Aug 2025 12:46:13
U	ICB14 - 1	ICB14	-0.0527	-140	PPB	SMPL	-7470A	29 Aug 2025 12:48:28
U	CCV40 - 1	CCV40	4.9709	20694	PPB	SMPL	-7470A	29 Aug 2025 12:50:43
U	CCB40 - 1	CCB40	-0.0614	-176	PPB	SMPL	-7470A	29 Aug 2025 12:52:58
U	CRA - 1	CRA	0.1966	894	PPB	SMPL	-7470A	29 Aug 2025 12:55:14
U	HighStd - 1	HighStd	9.7903	40681	PPB	SMPL	-7470A	29 Aug 2025 13:00:12
U	ChkStd - 1	ChkStd	7.2182	30014	PPB	SMPL	-7470A	29 Aug 2025 13:02:27
U	PB169473BL - 1	PBW	-0.0855	-276	PPB	SMPL	-7470A	29 Aug 2025 13:04:45
U	PB169473BS - 1	LCSW	4.3898	18284	PPB	SMPL	-7470A	29 Aug 2025 13:07:03
U	Q2971-37 - 1	GROUNDWATER	-0.0172	7	PPB	SMPL	-7470A	29 Aug 2025 13:09:19
U	Q2975-02 - 1	70925	0.1626	753	PPB	SMPL	-7470A	29 Aug 2025 13:11:37
U	Q2977-02 - 1	81925	0.9463	4003	PPB	SMPL	-7470A	29 Aug 2025 13:13:54
U	Q2977-03 - 1	82125	0.4436	1918	PPB	SMPL	-7470A	29 Aug 2025 13:16:19
U	Q2977-04 - 1	82225	0.1831	838	PPB	SMPL	-7470A	29 Aug 2025 13:18:36
U	CCV41 - 1	CCV41	4.6302	19281	PPB	SMPL	-7470A	29 Aug 2025 13:20:55
U	CCB41 - 1	CCB41	-0.0283	-39	PPB	SMPL	-7470A	29 Aug 2025 13:26:54
U	Q2977-05 - 1	82725	0.6876	2930	PPB	SMPL	-7470A	29 Aug 2025 13:36:00
U	Q2977-05DUP - 1	82725DUP	0.6321	2700	PPB	SMPL	-7470A	29 Aug 2025 13:38:15
U	Q2977-05MS - 1	82725MS	1.8725	7844	PPB	SMPL	-7470A	29 Aug 2025 13:40:32
U	Q2977-05MSD - 1	82725MSD	2.2110	9248	PPB	SMPL	-7470A	29 Aug 2025 13:42:51
U	Q2981-01 - 1	1127-COMP	-0.0768	-240	PPB	SMPL	-7470A	29 Aug 2025 13:45:10
U	Q2981-02 - 1	1203	0.5308	2280	PPB	SMPL	-7470A	29 Aug 2025 13:49:15
U	Q2977-05LX5 - 1		0.0676	359	PPB	SMPL	-7470A	29 Aug 2025 13:51:31
U	Q2977-05A - 1		3.1425	13111	PPB	SMPL	-7470A	29 Aug 2025 13:53:50
U	CCV42 - 1	CCV42	4.9526	20618	PPB	SMPL	-7470A	29 Aug 2025 14:01:59
U	CCB42 - 1	CCB42	-0.0811	-258	PPB	SMPL	-7470A	29 Aug 2025 14:04:14