

L3137028


7471B

INSTRUMENT ID: CV1

Linear

μ Abs.:

59767



A= 0.0000e+000

B= 1.6577e-004

C= 2.6703e-002

Rho= 0.9996747

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	C/D
0.0	0.000	0.026	0.026	-3	0.000	-3					-
0.2	0.200	0.231	0.031	1233	0.0 %	1233					16
2.5	2.500	2.453	-0.047	14637	0.0 %	14637					-2
5.0	5.000	4.880	-0.120	29277	0.0 %	29277					-2
7.5	7.500	7.675	0.175	46137	0.0 %	46137					2
10.0	10.000	9.934	-0.066	59767	0.0 %	59767					-1

LB137028 INSTRUMENT ID : CV1

Method: 7471B

Operator: Admin

Date of Analysis: 02 Sep 2025 09:22:07

Type	Sample ID	Extended ID	Conc.	μ Abs.	Units	Type	Std ConcMethod	Date
S	0.0 - 1	50	-	-3	PPB	Std	0.00007471B	02 Sep 2025 10:03:32
S	0.2 - 1	50.2	-	1233	PPB	Std	0.20007471B	02 Sep 2025 10:11:43
S	2.5 - 1	52.5	-	14637	PPB	Std	2.50007471B	02 Sep 2025 10:14:00
S	5.0 - 1	55	-	29277	PPB	Std	5.00007471B	02 Sep 2025 10:16:17
S	7.5 - 1	57.5	-	46137	PPB	Std	7.50007471B	02 Sep 2025 10:18:34
S	10.0 - 1	59	-	59767	PPB	Std	10.00007471B	02 Sep 2025 10:20:52
U	ICV15 - 1	ICV15	4.3746	26228	PPB	SMPL	-7471B	02 Sep 2025 10:29:20
U	ICB15 - 1	ICB15	0.0088	-108	PPB	SMPL	-7471B	02 Sep 2025 10:31:35
U	CCV43 - 1	CCV43	4.5278	27152	PPB	SMPL	-7471B	02 Sep 2025 10:33:52
U	CCB43 - 1	CCB43	-0.0069	-203	PPB	SMPL	-7471B	02 Sep 2025 10:36:07
U	CRA - 1	CRA	0.1961	1022	PPB	SMPL	-7471B	02 Sep 2025 10:38:25
U	HighStd - 1	HighStd	9.9887	60094	PPB	SMPL	-7471B	02 Sep 2025 10:40:40
U	ChkStd - 1	ChkStd	6.9521	41776	PPB	SMPL	-7471B	02 Sep 2025 10:42:55
U	PB169487BL - 1	PBS	-0.0459	-438	PPB	SMPL	-7471B	02 Sep 2025 10:45:14
U	PB169487BS - 1	LCSS	3.5676	21360	PPB	SMPL	-7471B	02 Sep 2025 10:50:13
U	Q2973-01 - 1	OK-01-082825	0.1567	784	PPB	SMPL	-7471B	02 Sep 2025 10:52:29
U	Q2979-01 - 1	SP-A	2.4679	14726	PPB	SMPL	-7471B	02 Sep 2025 10:54:47
U	Q2979-03 - 1	SP-B	8.2164	49403	PPB	SMPL	-7471B	02 Sep 2025 10:57:06
U	Q2986-01 - 1	NORTH-TP-1	0.1414	692	PPB	SMPL	-7471B	02 Sep 2025 10:59:25
U	Q2986-03 - 1	NORTH-TP-2	0.1565	783	PPB	SMPL	-7471B	02 Sep 2025 11:01:53
U	CCV44 - 1	CCV44	4.8575	29141	PPB	SMPL	-7471B	02 Sep 2025 11:06:57
U	CCB44 - 1	CCB44	0.0018	-150	PPB	SMPL	-7471B	02 Sep 2025 11:09:13
U	Q2986-05 - 1	NORTH-TP-3	0.1718	875	PPB	SMPL	-7471B	02 Sep 2025 11:11:31
U	Q2986-05DUP - 1	NORTH-TP-3DUP	0.1300	623	PPB	SMPL	-7471B	02 Sep 2025 11:13:54
U	Q2986-05MS - 1	NORTH-TP-3MS	3.7248	22308	PPB	SMPL	-7471B	02 Sep 2025 11:16:18
U	Q2986-05MSD - 1	NORTH-TP-3MSD	3.6091	21610	PPB	SMPL	-7471B	02 Sep 2025 11:18:33
U	Q2986-05LX5 - 1		0.0375	65	PPB	SMPL	-7471B	02 Sep 2025 11:23:51
U	Q2986-05A - 1		3.3243	19892	PPB	SMPL	-7471B	02 Sep 2025 11:26:32
U	CCV45 - 1	CCV45	5.0084	30051	PPB	SMPL	-7471B	02 Sep 2025 11:28:48
U	CCB45 - 1	CCB45	-0.0061	-198	PPB	SMPL	-7471B	02 Sep 2025 11:31:08