

Analytical Summary Report

Analysis Method: 9040C

Analyst By : jignesh

Parameter: pH

Supervisor Review By : Iwona

Run Number: LB137030

Slope : 98.6

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W3178
BUFFER PH 7.00 GREEN 1PINT PK6	W3093
PH 10.01 BUFFER,COLOR CD 475ML	W3191
buffer solution pH 7 yellow	W3217
Buffer Solution, PH2 (500ml)	W3161
pH 12.00 Buffer	W3200

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.05].

True Value of CCV2 = 12.00 Control Limits[+/- 0.05].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.2	4.01	09/02/2025	11:15
2	CAL2	1	Water	NA	NA	20.2	7.00	09/02/2025	11:16
3	CAL3	1	Water	NA	NA	20.3	10.02	09/02/2025	11:19
4	ICV	1	Water	NA	NA	20.2	6.99	09/02/2025	11:20
5	CCV1	1	Water	NA	NA	20.2	2.01	09/02/2025	11:22
6	Q2990-01	1	Water	NA	NA	23.4	8.01	09/02/2025	11:30
7	Q2990-01DUP	1	Water	NA	NA	23.5	8.02	09/02/2025	11:32
8	Q2990-02	1	Water	NA	NA	23.4	6.94	09/02/2025	11:37
9	Q2990-03	1	Water	NA	NA	23.2	9.72	09/02/2025	11:55
10	Q2990-04	1	Water	NA	NA	23.3	10.10	09/02/2025	12:00
11	CCV2	1	Water	NA	NA	20.2	12.02	09/02/2025	12:05

WORKLIST(Hardcopy Internal Chain)

U3B37030

WorkList Name : ph w q2990 WorkList ID : 191624 Department : Wet-Chemistry Date : 09-02-2025 11:02:00

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
Q2990-01	A 310	Water	pH	Cool 4 deg C	PSEG03	D31	08/29/2025	9040C
Q2990-02	A 308	Water	pH	Cool 4 deg C	PSEG03	D31	08/29/2025	9040C
Q2990-03	C 306	Water	pH	Cool 4 deg C	PSEG03	D31	08/29/2025	9040C
Q2990-04	C 307	Water	pH	Cool 4 deg C	PSEG03	D31	08/29/2025	9040C

Date/Time 09/02/25 11:11
 Raw Sample Received by: [Signature]
 Raw Sample Relinquished by: [Signature]

Date/Time 09/02/25 15:30
 Raw Sample Received by: [Signature]
 Raw Sample Relinquished by: [Signature]