

Analytical Summary Report

Analysis Method: 1010B

Reviewed By: Eman

Parameter: Flash Point

Supervisor Review By: Iwona

Run Number: LB137035

Ambient Barometric Pressure (mmHg): 760.00

Thermometer ID: Flashpoint

Barometric Scale ID: 0511064

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W3194

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	9	28.00	82.4	82.4	09/02/2025	12:00
2	Q2990-01		1	15	100.00	>212.0	>212.0	09/02/2025	12:30
3	Q2990-01DUP		1	15	100.00	>212.0	>212.0	09/02/2025	13:00
4	Q2990-02		1	9	86.00	186.8	186.8	09/02/2025	13:30
5	Q2990-03		1	13	100.00	>212.0	>212.0	09/02/2025	14:00
6	Q2990-04		1	11	100.00	>212.0	>212.0	09/02/2025	14:30

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06

WORKLIST(Hardcopy Internal Chain)

16137035

WorkList Name : fp-9-02

WorkList ID : 191601

Department : Wet-Chemistry

Date : 09-02-2025 08:05:55

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
Q2990-01	310	Water	Flash Point	Cool 4 deg C	PSEG03	D31	08/29/2025	1010B
Q2990-02	308	Water	Flash Point	Cool 4 deg C	PSEG03	D31	08/29/2025	1010B
Q2990-03	306	Water	Flash Point	Cool 4 deg C	PSEG03	D31	08/29/2025	1010B
Q2990-04	307	Water	Flash Point	Cool 4 deg C	PSEG03	D31	08/29/2025	1010B

Date/Time 09/02/25 11:45
Raw Sample Received by: EM(wc)
Raw Sample Relinquished by: J8(wc)

Date/Time 09/02/25 15:10
Raw Sample Received by: J8(wc)
Raw Sample Relinquished by: EM(wc)