

LB137040


7470A

INSTRUMENT ID: CV1

Linear

μ Abs.:

46222



A= 0.0000e+000

B= 2.1486e-004

C= 5.5647e-002

Rho= 0.9994577

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	0.061	0.061	23	0.000	23	0			
0.2	0.200	0.255	0.055	928	0.0 %	928	0			
2.5	2.500	2.452	-0.048	11155	0.0 %	11155	0			
5.0	5.000	4.775	-0.225	21965	0.0 %	21965				
7.5	7.500	7.670	0.170	35438	0.0 %	35438				
10.0	10.000	9.987	-0.013	46222	0.0 %	46222				

LB137040 INSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 02 Sep 2025 15:04:25

Sample ID	Extended ID	μ Abs	Conc	Std Conc Method	Units	Date	Type	Type	B	C	Rho
0.0 - 1	50	23	-	0.0007470A	PPB	02 Sep 2025 15:04:37	S	Std	-	-	-
0.2 - 1	50.2	928	-	0.20007470A	PPB	02 Sep 2025 15:06:54	S	Std	-	-	-
2.5 - 1	52.5	1115	-	2.50007470A	PPB	02 Sep 2025 15:09:11	S	Std	-	-	-
5.0 - 1	55	2196	-	5.00007470A	PPB	02 Sep 2025 15:11:28	S	Std	-	-	-
7.5 - 1	57	3543	-	7.50007470A	PPB	02 Sep 2025 15:13:45	S	Std	-	-	-
10.0 - 1	57	0	-	10.0007470A	PPB	02 Sep 2025 15:18:56	S	Std	-	-	-
ICV17 - 1	ICV17	19343	4.2547	-7470A	PPB	02 Sep 2025 15:25:59	U	SMPL	2.14861e-004	5.56469e-002	9.99458e-001
ICB17 - 1	ICB17	-136	0.0264	-7470A	PPB	02 Sep 2025 15:28:13	U	SMPL	2.14861e-004	5.56469e-002	9.99458e-001
CCV48 - 1	CCV48	23115	5.0222	-7470A	PPB	02 Sep 2025 15:30:28	U	SMPL	2.14861e-004	5.56469e-002	9.99458e-001
CCB48 - 1	CCB48	-157	0.0219	-7470A	PPB	02 Sep 2025 15:32:43	U	SMPL	2.14861e-004	5.56469e-002	9.99458e-001
CRA - 1	CRA	909	0.2510	-7470A	PPB	02 Sep 2025 15:34:59	U	SMPL	2.14861e-004	5.56469e-002	9.99458e-001
HighStd - 1	HighStd	44620	9.6428	-7470A	PPB	02 Sep 2025 15:37:14	U	SMPL	2.14861e-004	5.56469e-002	9.99458e-001
ChkStd - 1	ChkStd	28443	6.1669	-7470A	PPB	02 Sep 2025 15:39:29	U	SMPL	2.14861e-004	5.56469e-002	9.99458e-001
PB169510BL - 1	PBW	-241	0.0039	-7470A	PPB	02 Sep 2025 15:41:46	U	SMPL	2.14861e-004	5.56469e-002	9.99458e-001
PB169510BS - 1	LCWSW	18612	4.0546	-7470A	PPB	02 Sep 2025 15:44:04	U	SMPL	2.14861e-004	5.56469e-002	9.99458e-001
Q2970-06 - 1	TP10	-144	0.0247	-7470A	PPB	02 Sep 2025 15:46:20	U	SMPL	2.14861e-004	5.56469e-002	9.99458e-001
Q2970-12 - 1	TP11	-6	0.0544	-7470A	PPB	02 Sep 2025 15:48:37	U	SMPL	2.14861e-004	5.56469e-002	9.99458e-001
Q2970-18 - 1	TP12	14	0.0587	-7470A	PPB	02 Sep 2025 15:50:54	U	SMPL	2.14861e-004	5.56469e-002	9.99458e-001
Q2970-24 - 1	TP9	17	0.0593	-7470A	PPB	02 Sep 2025 15:53:11	U	SMPL	2.14861e-004	5.56469e-002	9.99458e-001
Q2970-30 - 1	TP6	25	0.0610	-7470A	PPB	02 Sep 2025 15:55:26	U	SMPL	2.14861e-004	5.56469e-002	9.99458e-001
CCV49 - 1	CCV49	23456	5.0953	-7470A	PPB	02 Sep 2025 16:05:51	U	SMPL	2.14861e-004	5.56469e-002	9.99458e-001
CCB49 - 1	CCB49	-177	0.0176	-7470A	PPB	02 Sep 2025 16:08:06	U	SMPL	2.14861e-004	5.56469e-002	9.99458e-001
Q2970-36 - 1	TP3	0	0.0556	-7470A	PPB	02 Sep 2025 16:10:23	U	SMPL	2.14861e-004	5.56469e-002	9.99458e-001
Q2971-06 - 1	TP2	22	0.0604	-7470A	PPB	02 Sep 2025 16:12:38	U	SMPL	2.14861e-004	5.56469e-002	9.99458e-001
Q2971-12 - 1	TP1	7	0.0572	-7470A	PPB	02 Sep 2025 16:14:53	U	SMPL	2.14861e-004	5.56469e-002	9.99458e-001
Q2971-18 - 1	TP5	21	0.0602	-7470A	PPB	02 Sep 2025 16:17:09	U	SMPL	2.14861e-004	5.56469e-002	9.99458e-001
Q2971-24 - 1	TP8	36	0.0634	-7470A	PPB	02 Sep 2025 16:19:25	U	SMPL	2.14861e-004	5.56469e-002	9.99458e-001
Q2971-30 - 1	TP7	21	0.0602	-7470A	PPB	02 Sep 2025 16:21:41	U	SMPL	2.14861e-004	5.56469e-002	9.99458e-001
Q2971-36 - 1	TP4	17	0.0593	-7470A	PPB	02 Sep 2025 16:23:57	U	SMPL	2.14861e-004	5.56469e-002	9.99458e-001
Q2982-02 - 1	MOO-25-0237	20	0.0599	-7470A	PPB	02 Sep 2025 16:26:14	U	SMPL	2.14861e-004	5.56469e-002	9.99458e-001
Q2982-04 - 1	MOO-25-0244-0248	32	0.0625	-7470A	PPB	02 Sep 2025 16:28:31	U	SMPL	2.14861e-004	5.56469e-002	9.99458e-001
Q2982-06 - 1	MOO-25-0249-0253	30	0.0621	-7470A	PPB	02 Sep 2025 16:30:47	U	SMPL	2.14861e-004	5.56469e-002	9.99458e-001
CCV50 - 1	CCV50	21810	4.7418	-7470A	PPB	02 Sep 2025 16:36:13	U	SMPL	2.14861e-004	5.56469e-002	9.99458e-001
CCB50 - 1	CCB50	-1	0.0554	-7470A	PPB	02 Sep 2025 16:38:28	U	SMPL	2.14861e-004	5.56469e-002	9.99458e-001
Q2983-02 - 1	LAW-25-0128	-31	0.0490	-7470A	PPB	02 Sep 2025 16:40:45	U	SMPL	2.14861e-004	5.56469e-002	9.99458e-001
Q2970-06dDUP - 1		20	0.0599	-7470A	PPB	02 Sep 2025 16:43:06	U	SMPL	2.14861e-004	5.56469e-002	9.99458e-001
Q2970-06MS - 1		20392	4.4371	-7470A	PPB	02 Sep 2025 16:45:23	U	SMPL	2.14861e-004	5.56469e-002	9.99458e-001
Q2970-06MSD - 1		20566	4.4728	-7470A	PPB	02 Sep 2025 16:47:39	U	SMPL	2.14861e-004	5.56469e-002	9.99458e-001
PB169428TB - 1	PB169428TB	-164	0.0204	-7470A	PPB	02 Sep 2025 16:48:56	U	SMPL	2.14861e-004	5.56469e-002	9.99458e-001
Q2970-06x5 - 1		48	0.0660	-7470A	PPB	02 Sep 2025 16:52:12	U	SMPL	2.14861e-004	5.56469e-002	9.99458e-001
Q2970-06a - 1		15269	3.3364	-7470A	PPB	02 Sep 2025 16:54:28	U	SMPL	2.14861e-004	5.56469e-002	9.99458e-001
CCV51 - 1	CCV51	22763	4.9465	-7470A	PPB	02 Sep 2025 16:56:45	U	SMPL	2.14861e-004	5.56469e-002	9.99458e-001
CCB51 - 1	CCB51	164	0.0887	-7470A	PPB	02 Sep 2025 16:59:03	U	SMPL	2.14861e-004	5.56469e-002	9.99458e-001