

Analytical Summary Report

Analysis Method: 1010B

Reviewed By: rubina

Parameter: Flash Point

Supervisor Review By: Iwona

Run Number: LB137061

Ambient Barometric Pressure (mmHg): 755.00

Thermometer ID: Flash Point

Barometric Scale ID: 0511064

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W3194

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	9	28.00	82.4	82.7	09/03/2025	13:40
2	Q3004-04		1	14	100.00	>212.0	>212.0	09/03/2025	14:10
3	Q3004-04DUP		1	14	100.00	>212.0	>212.0	09/03/2025	14:40

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06

WORKLIST(Hardcopy Internal Chain)

L6137061

WorkList Name : FP-9-03

WorkList ID : 191652

Department : Wet-Chemistry

Date : 09-03-2025 10:45:25

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
Q3004-04	DRUM-OILY-WATER	Water	Flash Point	Cool 4 deg C	PSEG03	J13	09/02/2025	1010B

Date/Time 09/03/2025 13:30
 Raw Sample Received by: RFLWLS
 Raw Sample Relinquished by: RFLWLS

Date/Time 09/03/2025 15:20
 Raw Sample Received by: RFLWLS
 Raw Sample Relinquished by: RFLWLS