



PERCENT SOLID

Supervisor: Iwona
Analyst: JIGNESH
Date: 9/5/2025

OVENTEMP IN Celsius(°C): 107
Time IN: 17:30
In Date: 09/04/2025
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
OvenID: M OVEN#1

OVENTEMP OUT Celsius(°C): 104
Time OUT: 08:33
Out Date: 09/05/2025
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
BalanceID: M SC-4
Thermometer ID: % SOLID-OVEN

QC:LB137064

Lab ID	Client SampleID	Dish #	Dish Wt(g) (A)	Sample Wt(g)	Dish + Sample Wt(g) (B)	Dish+Dry Sample Wt(g) (C)	% Solid	Comments
Q3011-01	WASTE	1	1.15	10.35	11.5	10.87	93.9	
Q3011-02	VOC	2	1.16	10.41	11.57	10.9	93.6	
Q3011-03	1	3	1.16	10.43	11.59	10.77	92.1	
Q3011-04	2	4	1.17	10.79	11.96	10.8	89.2	
Q3011-05	3	5	1.15	10.33	11.48	10.73	92.7	
Q3011-06	4	6	1.13	10.33	11.46	10.79	93.5	
Q3011-07	5	7	1.19	10.43	11.62	11.22	96.2	
Q3011-08	6	8	1.19	10.40	11.59	10.81	92.5	
Q3012-01	WASTE	9	1.16	10.78	11.94	11.34	94.4	
Q3012-02	VOC	10	1.14	10.82	11.96	11.48	95.6	
Q3012-03	1	11	1.15	10.30	11.45	10.86	94.3	
Q3012-04	2	12	1.17	10.77	11.94	11.22	93.3	
Q3012-05	3	13	1.15	10.81	11.96	11.31	94.0	
Q3012-06	4	14	1.16	10.39	11.55	11.00	94.7	
Q3012-07	5	15	1.12	10.19	11.31	10.5	92.1	
Q3012-07D	5DUP	16	1.17	10.17	11.34	10.53	92.0	
Q3014-01	Red Firestop Sealant	17	1.00	1.00	2.00	2.00	100.0	Caulk sample
Q3014-02	Sidewalk Joint Caulk	18	1.00	1.00	2.00	2.00	100.0	Caulk sample
Q3014-03	Ext Metal Wall panel Caulk	19	1.00	1.00	2.00	2.00	100.0	Caulk sample
Q3021-02	821-ABC	20	1.00	1.00	2.00	2.00	100.0	oil sample
Q3021-04	821-DEF	21	1.00	1.00	2.00	2.00	100.0	oil sample
Q3021-06	821-GHI	22	1.00	1.00	2.00	2.00	100.0	oil sample
Q3023-01	TP-1	23	1.14	10.62	11.76	10.5	88.1	
Q3023-04	TP-2	24	1.19	10.51	11.7	10.9	92.4	
Q3026-01	EO-02-09042025	25	1.14	10.85	11.99	11.08	91.6	
Q3026-02	EO-02-09042025-E2	26	1.18	10.36	11.54	10.56	90.5	
Q3027-01	HD-01-09042025	27	1.11	10.55	11.66	11.06	94.3	
Q3027-02	HD-01-09042025-E2	28	1.19	10.30	11.49	11.05	95.7	

$$\% \text{ Solid} = \frac{(C-A) * 100}{(B-A)}$$

WORKLIST(Hardcopy Internal Chain)

1370625

WorkList Name : %1-090425 WorkList ID : 191662 Department : Wet-Chemistry Date : 09-04-2025 09:55:58

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
Q3011-01	WASTE	Solid	Percent Solids	Cool 4 deg C	SCIA01	J33	09/03/2025	Chemtech -SO
Q3011-02	VOC	Solid	Percent Solids	Cool 4 deg C	SCIA01	J33	09/03/2025	Chemtech -SO
Q3011-03	1	Solid	Percent Solids	Cool 4 deg C	SCIA01	J33	09/03/2025	Chemtech -SO
Q3011-04	2	Solid	Percent Solids	Cool 4 deg C	SCIA01	J33	09/03/2025	Chemtech -SO
Q3011-05	3	Solid	Percent Solids	Cool 4 deg C	SCIA01	J33	09/03/2025	Chemtech -SO
Q3011-06	4	Solid	Percent Solids	Cool 4 deg C	SCIA01	J33	09/03/2025	Chemtech -SO
Q3011-07	5	Solid	Percent Solids	Cool 4 deg C	SCIA01	J33	09/03/2025	Chemtech -SO
Q3011-08	6	Solid	Percent Solids	Cool 4 deg C	SCIA01	J33	09/03/2025	Chemtech -SO
Q3012-01	WASTE	Solid	Percent Solids	Cool 4 deg C	SCIA01	J33	09/03/2025	Chemtech -SO
Q3012-02	VOC	Solid	Percent Solids	Cool 4 deg C	SCIA01	J42	09/03/2025	Chemtech -SO
Q3012-03	1	Solid	Percent Solids	Cool 4 deg C	SCIA01	J42	09/03/2025	Chemtech -SO
Q3012-04	2	Solid	Percent Solids	Cool 4 deg C	SCIA01	J42	09/03/2025	Chemtech -SO
Q3012-05	3	Solid	Percent Solids	Cool 4 deg C	SCIA01	J42	09/03/2025	Chemtech -SO
Q3012-06	4	Solid	Percent Solids	Cool 4 deg C	SCIA01	J42	09/03/2025	Chemtech -SO
Q3012-07	5	Solid	Percent Solids	Cool 4 deg C	SCIA01	J42	09/03/2025	Chemtech -SO
Q3014-01	Red Firestop Sealant	Solid	Percent Solids	Cool 4 deg C	ATCE02	J11	09/03/2025	Chemtech -SO
Q3014-02	Sidewalk Joint Caulk	Solid	Percent Solids	Cool 4 deg C	ATCE02	J11	09/03/2025	Chemtech -SO
Q3014-03	Ext Metal Wall panel Caulk	Solid	Percent Solids	Cool 4 deg C	ATCE02	J11	09/03/2025	Chemtech -SO
Q3021-02	821-ABC	Solid	Percent Solids	Cool 4 deg C	PSEG03	F31	09/04/2025	Chemtech -SO
Q3021-04	821-DEF	Solid	Percent Solids	Cool 4 deg C	PSEG03	F31	09/04/2025	Chemtech -SO
Q3021-06	821-GHI	Solid	Percent Solids	Cool 4 deg C	PSEG03	F31	09/04/2025	Chemtech -SO

Date/Time 09/04/25 16:50
 Raw Sample Received by: Ed Cady
 Raw Sample Relinquished by: Ed Cady
 Date/Time 09/04/25 18:00
 Raw Sample Received by: Ed Cady
 Raw Sample Relinquished by: Ed Cady

WORKLIST(Hardcopy Internal Chain)

12/25/2024

WorkList Name : %1-090425 WorkList ID : 191662 Department : Wet-Chemistry Date : 09-04-2025 09:55:58

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
Q3023-01	TP-1	Solid	Percent Solids	Cool 4 deg C	PSEG03	D41	09/04/2025	Chemtech -SO
Q3023-04	TP-2	Solid	Percent Solids	Cool 4 deg C	PSEG03	D41	09/04/2025	Chemtech -SO
Q3026-01	EO-02-09042025	Solid	Percent Solids	Cool 4 deg C	PSEG05	D21	09/04/2025	Chemtech -SO
Q3026-02	EO-02-09042025-E2	Solid	Percent Solids	Cool 4 deg C	PSEG05	D21	09/04/2025	Chemtech -SO
Q3027-01	HD-01-09042025	Solid	Percent Solids	Cool 4 deg C	PSEG05	D21	09/04/2025	Chemtech -SO
Q3027-02	HD-01-09042025-E2	Solid	Percent Solids	Cool 4 deg C	PSEG05	D21	09/04/2025	Chemtech -SO

Date/Time 09/04/25 16:50
 Raw Sample Received by: SP WOC
 Raw Sample Relinquished by: CP SM

Date/Time 09/04/25 18:00
 Raw Sample Received by: CP SM
 Raw Sample Relinquished by: SP WOC