

66137067

Test results

Aquakem 7.2AQ1

Page: 1

Alliance Technical Group

284 Sheffield Street, Mountainside, NJ 07092

Reviewed by : RM

Instrument ID : Konelab

9/4/2025 11:35

Test: CNEPA-NEW

Sample Id	Result	Dil. 1 +	Response	Errors
ICV001 ICV001	94.153	0.0	0.091	
ICB001 ICB001	0.657	0.0	0.002	
CCV001 CCV001	241.759	0.0	0.231	
CCB001 CCB001	0.805	0.0	0.002	
PB169537BL PBS537	0.460	0.0	0.002	
02899 04-MDL-SOIL-QT3-2025-06	0.956	0.0	0.005	
PB169538BL PBW538	0.565	0.0	0.002	
02899 02-MDL-WATER-QT3-2025-06	4.262	0.0	0.005	
09/04/2025 RM CCV002 CCV002	246.603	0.0	0.236	
CCB002 CCB002	0.820	0.0	0.002	

N	10
Mean	59.404
SD	101.5834
CV%	171.00

Aquakem v. 7.2AQ1

Results from time period:

Thu Sep 04 10:46:19 2025

Thu Sep 04 11:33:43 2025

Sample Id	Sar Test short r Tes Result	Result unit	Result date and time Stat
S0.0	A CNEPA-NE\ P 0.7774	µg/l	9/4/2025 10:46:19
S5.0	A CNEPA-NE\ P 4.688	µg/l	9/4/2025 10:46:20
S10.0	A CNEPA-NE\ P 9.8203	µg/l	9/4/2025 10:46:21
S100.0	A CNEPA-NE\ P 100.6659	µg/l	9/4/2025 10:46:22
S250.0	A CNEPA-NE\ P 248.3497	µg/l	9/4/2025 10:46:23
S500.0	A CNEPA-NE\ P 500.6986	µg/l	9/4/2025 10:46:24
ICV001 ICV001	S CNEPA-NE\ P 94.1526	µg/l	9/4/2025 11:26:57
ICB001 ICB001	S CNEPA-NE\ P 0.6571	µg/l	9/4/2025 11:26:58
CCV001 CCV001	S CNEPA-NE\ P 241.7588	µg/l	9/4/2025 11:27:01
CCB001 CCB001	S CNEPA-NE\ P 0.805	µg/l	9/4/2025 11:27:02
PB169537BL PBS537	S CNEPA-NE\ P 0.4604	µg/l	9/4/2025 11:27:05
Q2899-04-MDL-SOIL-QT3-2025-06	S CNEPA-NE\ P 3.9563	µg/l	9/4/2025 11:33:35
PB169538BL PBW538	S CNEPA-NE\ P 0.5653	µg/l	9/4/2025 11:33:37
Q2899-02-MDL-WATER-QT3-2025-06	S CNEPA-NE\ P 4.2619	µg/l	9/4/2025 11:33:38
CCV002 CCV002	S CNEPA-NE\ P 246.6027	µg/l	9/4/2025 11:33:41
CCB002 CCB002	S CNEPA-NE\ P 0.8199	µg/l	9/4/2025 11:33:43

Alliance Technical Group
284 Sheffield Street, Mountainside, NJ 07092

Reviewed by : RM Instrument ID : Konelab

9/4/2025 10:47


Test CNEPA-NEW

Accepted 9/4/2025 10:47

Factor Slope 1049 0.000953
Bias y-intercept 0.001 09/04/2025
RM

Coeff. of det. 0.999978

Errors



	Calibrator	Response	Calc. con.	Conc.	Errors
1	0.0PPBCN 50.0	0.002	0.7774	0.0000	-6.2
2	5.0PPBCN 55.0	0.006	4.6880	5.0000	-1.8
3	10PPBCN 10.0	0.010	9.8203	10.0000	0.7
4	100PPBCN 100.0	0.097	100.6659	100.0000	-0.7
5	250PPBCN 250.0	0.238	248.3497	250.0000	0.1
6	500PPBCN 500.0	0.478	500.6986	500.0000	

09/04/2025
RM