

Analytical Summary Report

Analysis Method: 9045D

Analyst By : JIGNESH

Parameter: Corrosivity

Supervisor Review By : Iwona

Run Number: LB137080

Slope : 98.5

BalanceID: WC SC-7

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W3178
BUFFER PH 7.00 GREEN 1PINT PK6	W3093
PH 10.01 BUFFER,COLOR CD 475ML	W3191
buffer solution pH 7 yellow	W3217
Buffer Solution, PH2 (500ml)	W3161
pH 12.00 Buffer	W3200

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.05].

True Value of CCV2 = 12.00 Control Limits[+/- 0.05].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.2	4.02	09/05/2025	09:00
2	CAL2	1	Water	NA	NA	20.2	7.01	09/05/2025	09:01
3	CAL3	1	Water	NA	NA	20.2	10.02	09/05/2025	09:03
4	ICV	1	Water	NA	NA	20.3	7.02	09/05/2025	09:05
5	CCV1	1	Water	NA	NA	20.2	2.01	09/05/2025	09:10
6	Q3023-02	1	Solid	20.02	20	23.2	8.01	09/05/2025	09:30
7	Q3023-02DUP	1	Solid	20.04	20	23.4	8.02	09/05/2025	09:33
8	Q3023-05	1	Solid	20.02	20	23.6	9.01	09/05/2025	09:44
9	CCV2	1	Water	NA	NA	20.3	12.02	09/05/2025	09:47

WORKLIST(Hardcopy Internal Chain)

VB 137080

WorkList Name : corrosivity q3023

WorkList ID : 191686

Department : Wet-Chemistry

Date : 09-05-2025 07:53:58

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
Q3023-02	TP-1	Solid	Corrosivity	Cool 4 deg C	PSEG03	D31	09/04/2025	9045D
Q3023-05	TP-2	Solid	Corrosivity	Cool 4 deg C	PSEG03	D41	09/04/2025	9045D

Date/Time

09/05/25 08:00

Raw Sample Received by:

sf woc

Raw Sample Relinquished by:

sf sm

Date/Time

09/05/25 10:30

Raw Sample Received by:

sf sm

Raw Sample Relinquished by:

sf woc