

Analytical Summary Report

Analysis Method: 9040C
Parameter: pH
Run Number: LB137081

Analyst By : JIGNESH

Supervisor Review By : Iwona

Slope : 99.2

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W3178
BUFFER PH 7.00 GREEN 1PINT PK6	W3093
PH 10.01 BUFFER,COLOR CD 475ML	W3191
buffer solution pH 7 yellow	W3217
Buffer Solution, PH2 (500ml)	W3161
pH 12.00 Buffer	W3200

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.05].

True Value of CCV2 = 12.00 Control Limits[+/- 0.05].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.2	4.01	09/05/2025	08:35
2	CAL2	1	Water	NA	NA	20.2	7.01	09/05/2025	08:36
3	CAL3	1	Water	NA	NA	20.3	10.02	09/05/2025	08:38
4	ICV	1	Water	NA	NA	20.2	7.01	09/05/2025	08:40
5	CCV1	1	Water	NA	NA	20.3	2.01	09/05/2025	08:44
6	Q3017-02	1	Water	NA	NA	20.6	7.25	09/05/2025	08:55
7	Q3017-02DUP	1	Water	NA	NA	20.7	7.27	09/05/2025	08:59
8	Q3017-03	1	Water	NA	NA	20.1	7.18	09/05/2025	09:00
9	Q3017-03DUP	1	Water	NA	NA	20.2	7.19	09/05/2025	09:02
10	Q3017-04	1	Water	NA	NA	20.4	6.83	09/05/2025	09:10
11	Q3017-04DUP	1	Water	NA	NA	20.6	6.86	09/05/2025	09:15
12	Q3017-05	1	Water	NA	NA	20.3	6.63	09/05/2025	09:25
13	Q3017-05DUP	1	Water	NA	NA	20.5	6.65	09/05/2025	09:30
14	Q3022-01	1	Water	NA	NA	23.6	7.43	09/05/2025	09:38
15	Q3022-01DUP	1	Water	NA	NA	23.8	7.45	09/05/2025	09:40
16	CCV2	1	Water	NA	NA	20.3	12.02	09/05/2025	09:42

WORKLIST(Hardcopy Internal Chain)

180137081

WorkList Name : ph q3022

WorkList ID : 191685

Department : Wet-Chemistry

Date : 09-05-2025 07:52:12

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
Q3017-02	TT-076-IDWGW-20250903	Water	pH	Cool 4 deg C	TETR06	J12	09/04/2025	9040C
Q3017-03	TT-077-IDWGW-20250903	Water	pH	Cool 4 deg C	TETR06	J12	09/04/2025	9040C
Q3017-04	TT-078-IDWGW-20250903	Water	pH	Cool 4 deg C	TETR06	J12	09/04/2025	9040C
Q3017-05	TT-079-IDWGW-20250903	Water	pH	Cool 4 deg C	TETR06	J12	09/04/2025	9040C
Q3022-01	I 1413	Water	pH	Cool 4 deg C	PSEG03	D31	09/04/2025	9040C

09/05/25 08:10

Date/Time

Raw Sample Received by:

JA WOC

Raw Sample Relinquished by:

CP SM

Date/Time 09/05/25 11:45

Raw Sample Received by:

CP SM

Raw Sample Relinquished by:

JA WOC