


LB137092
INSTRUMENT ID: CV,

245.1

μ Abs.:

44877



Linear

A= 0.0000e+000

B= 2.2202e-004

C= 1.0364e-002

Rho= 0.9999765

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	6/01
0.0	0.000	0.001	0.001	-40	0.000	-40					-
0.05	0.050	0.058	0.008	214	0.0 %	214					16
0.2	0.200	0.207	0.007	886	0.0 %	886					3
2.5	2.500	2.454	-0.046	11008	0.0 %	11008					-2
5.0	5.000	5.019	0.019	22557	0.0 %	22557					0
7.5	7.500	7.537	0.037	33898	0.0 %	33898					0
10.0	10.000	9.974	-0.026	44877	0.0 %	44877					0

LB137091 INSTRUMENT ID CV:1

Method: 245.1

Operator: Admin

Date of Analysis: 05 Sep 2025 11:53:45

Type	Sample ID	Extended ID	Conc.	µ Abs.	Units	Type	Std ConcMethod	Date
S	0.0 - 1	50	-	-40	PPB	Std	0.0000245.1	05 Sep 2025 12:10:35
S	0.05 - 1	90.05	-	214	PPB	Std	0.0500245.1	05 Sep 2025 12:13:45
S	0.2 - 1	50.2	-	886	PPB	Std	0.2000245.1	05 Sep 2025 12:16:02
S	2.5 - 1	52.5	-	11008	PPB	Std	2.5000245.1	05 Sep 2025 12:18:18
S	5.0 - 1	55	-	22557	PPB	Std	5.0000245.1	05 Sep 2025 12:20:35
S	7.5 - 1	59.5	-	33898	PPB	Std	7.5000245.1	05 Sep 2025 12:22:51
S	10.0 - 1	51.4	-	44877	PPB	Std	10.0000245.1	05 Sep 2025 12:25:09
U	ICV20 - 1	ICV20	3.9331	17668	PPB	SMPL	-245.1	05 Sep 2025 12:28:25
U	ICB20 - 1	ICB20	-0.0527	-284	PPB	SMPL	-245.1	05 Sep 2025 12:30:41
U	CCV58 - 1	CCV58	5.0414	22660	PPB	SMPL	-245.1	05 Sep 2025 12:32:58
U	CCB58 - 1	CCB58	-0.0469	-258	PPB	SMPL	-245.1	05 Sep 2025 12:35:13
U	CRA - 1	CRA	0.0519	187	PPB	SMPL	-245.1	05 Sep 2025 12:37:30
U	HighStd - 1	HighStd	9.9714	44865	PPB	SMPL	-245.1	05 Sep 2025 12:39:45
U	ChkStd - 1	ChkStd	6.8041	30599	PPB	SMPL	-245.1	05 Sep 2025 12:42:01
U	PB169580BL - 1	PBW	-0.0833	-422	PPB	SMPL	-245.1	05 Sep 2025 12:44:18
U	PB169580BS - 1	LCSW	3.9921	17934	PPB	SMPL	-245.1	05 Sep 2025 12:46:36
U	Q2126-10 - 1		0.0548	200	PPB	SMPL	-245.1	05 Sep 2025 12:51:33
U	Q2126-10RE - 1		0.1671	706	PPB	SMPL	-245.1	05 Sep 2025 12:53:49
U	Q2126-12 - 1	MDL-WATER-06-QT2-2025	0.0599	223	PPB	SMPL	-245.1	05 Sep 2025 13:07:26
U	CCV59 - 1	CCV59	5.0521	22708	PPB	SMPL	-245.1	05 Sep 2025 13:09:43
U	CCB59 - 1	CCB59	-0.0378	-217	PPB	SMPL	-245.1	05 Sep 2025 13:11:58