

LB137094


7470A

INSTRUMENT ID: CV1

Linear

 μ Abs.:

46438



A= 0.0000e+000

B= 2.1888e-004

C= 1.9437e-003

Rho= 0.9994094

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%D
0.0	0.000	0.018	0.018	72	0.000	72	0				—
0.2	0.200	0.199	-0.001	899	0.0 %	899	0				-1
2.5	2.500	2.605	0.105	11894	0.0 %	11894	0				4
5.0	5.000	4.945	-0.055	22583	0.0 %	22583					-1
7.5	7.500	7.267	-0.233	33191	0.0 %	33191					-3
10.0	10.000	10.166	0.166	46438	0.0 %	46438					2

LB137094 INSTRUMENT ID CV:1

Method: 7470A

Operator: Admin

Date of Analysis: 05 Sep 2025 14:04:28

Type	Sample ID	Extended ID	Conc	μ Abs	Units	Type	Std Conc	Method	Date
S	0.0 - 1	50	-	72	PPB	Std	0.0000	7470A	05 Sep 2025 14:42:29
S	0.2 - 1	50.2	-	899	PPB	Std	0.2000	7470A	05 Sep 2025 14:49:46
S	2.5 - 1	52.5	-	11894	PPB	Std	2.5000	7470A	05 Sep 2025 14:52:03
S	5.0 - 1	55	-	22583	PPB	Std	5.0000	7470A	05 Sep 2025 14:54:20
S	7.5 - 1	57.5	-	33191	PPB	Std	7.5000	7470A	05 Sep 2025 14:56:37
S	10.0 - 1	59.5	-	46438	PPB	Std	10.0000	7470A	05 Sep 2025 15:01:32
U	ICV21 - 1	ICV21	3.8173	16817	PPB	SMPL	-7470A	-7470A	05 Sep 2025 15:20:48
U	ICB21 - 1	ICB21	0.0454	75	PPB	SMPL	-7470A	-7470A	05 Sep 2025 15:23:10
U	CCV60 - 1	CCV60	4.9699	21933	PPB	SMPL	-7470A	-7470A	05 Sep 2025 15:25:27
U	CCB60 - 1	CCB60	0.0026	-115	PPB	SMPL	-7470A	-7470A	05 Sep 2025 15:27:42
U	CRA - 1	CRA	0.2288	889	PPB	SMPL	-7470A	-7470A	05 Sep 2025 15:29:59
U	HighStd - 1	HighStd	10.4556	46282	PPB	SMPL	-7470A	-7470A	05 Sep 2025 15:32:14
U	ChkStd - 1	ChkStd	6.9946	30920	PPB	SMPL	-7470A	-7470A	05 Sep 2025 15:37:04
U	PB169582BL - 1	PBW	-0.0065	-155	PPB	SMPL	-7470A	-7470A	05 Sep 2025 15:39:19
U	PB169582BS - 1	LCSW	3.8013	16746	PPB	SMPL	-7470A	-7470A	05 Sep 2025 15:41:38
U	Q3017-02 - 1	TT-076-IDWGW-20250903	0.0463	79	PPB	SMPL	-7470A	-7470A	05 Sep 2025 15:43:54
U	Q3017-03 - 1	TT-077-IDWGW-20250903	0.1093	359	PPB	SMPL	-7470A	-7470A	05 Sep 2025 15:46:13
U	Q3017-04 - 1	TT-078-IDWGW-20250903	0.0402	52	PPB	SMPL	-7470A	-7470A	05 Sep 2025 15:48:30
U	Q3017-05 - 1	TT-079-IDWGW-20250903	0.0235	-22	PPB	SMPL	-7470A	-7470A	05 Sep 2025 15:50:47
U	Q3017-05DUP - 1	TT-079-IDWGW-20250903DUP	0.0535	111	PPB	SMPL	-7470A	-7470A	05 Sep 2025 15:53:02
U	CCV61 - 1	CCV61	5.0386	22238	PPB	SMPL	-7470A	-7470A	05 Sep 2025 15:55:17
U	CCB61 - 1	CCB61	0.0008	-123	PPB	SMPL	-7470A	-7470A	05 Sep 2025 15:57:32
U	Q3017-05MS - 1	TT-079-IDWGW-20250903MS	3.7287	16424	PPB	SMPL	-7470A	-7470A	05 Sep 2025 16:02:42
U	Q3017-05MSD - 1	TT-079-IDWGW-20250903MSD	3.7204	16387	PPB	SMPL	-7470A	-7470A	05 Sep 2025 16:04:57
U	Q3022-01 - 1	1413	-0.0006	-129	PPB	SMPL	-7470A	-7470A	05 Sep 2025 16:07:17
U	PB169585BL - 1	PBW	0.0120	-73	PPB	SMPL	-7470A	-7470A	05 Sep 2025 16:09:33
U	PB169585BS - 1	LCSW	3.7963	16724	PPB	SMPL	-7470A	-7470A	05 Sep 2025 16:11:49
U	Q3023-02 - 1	TP-1	-0.0227	-227	PPB	SMPL	-7470A	-7470A	05 Sep 2025 16:14:05
U	Q3023-05 - 1	TP-2	0.0325	18	PPB	SMPL	-7470A	-7470A	05 Sep 2025 16:16:21
U	Q3023-05DUP - 1	TP-2DUP	0.0346	27	PPB	SMPL	-7470A	-7470A	05 Sep 2025 16:18:39
U	Q3023-05MS - 1	TP-2MS	3.7411	16479	PPB	SMPL	-7470A	-7470A	05 Sep 2025 16:20:55
U	Q3023-05MSD - 1	TP-2MSD	3.7296	16428	PPB	SMPL	-7470A	-7470A	05 Sep 2025 16:23:11
U	CCV62 - 1	CCV62	4.9724	21944	PPB	SMPL	-7470A	-7470A	05 Sep 2025 16:25:26
U	CCB62 - 1	CCB62	0.0073	-94	PPB	SMPL	-7470A	-7470A	05 Sep 2025 16:27:47
U	Q3017-05LX5 - 1		0.0442	70	PPB	SMPL	-7470A	-7470A	05 Sep 2025 16:30:06
U	Q3017-05A - 1		4.0338	17778	PPB	SMPL	-7470A	-7470A	05 Sep 2025 16:32:22
U	PB169556TB - 1	PB169556TB	-0.0362	-287	PPB	SMPL	-7470A	-7470A	05 Sep 2025 16:34:37
U	Q3023-05LX5 - 1		0.0226	-26	PPB	SMPL	-7470A	-7470A	05 Sep 2025 16:36:55
U	Q3023-05A - 1		4.0444	17825	PPB	SMPL	-7470A	-7470A	05 Sep 2025 16:39:11
U	CCV63 - 1	CCV63	5.0742	22396	PPB	SMPL	-7470A	-7470A	05 Sep 2025 16:41:27
U	CCV63 - 1	CCB63	-0.0024	-137	PPB	SMPL	-7470A	-7470A	05 Sep 2025 16:43:46