

Analytical Summary Report

Analysis Method: 9040C

Analyst By : JIGNESH

Parameter: pH

Supervisor Review By : Iwona

Run Number: LB137101

Slope : 98.5

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W3178
BUFFER PH 7.00 GREEN 1PINT PK6	W3093
PH 10.01 BUFFER,COLOR CD 475ML	W3191
buffer solution pH 7 yellow	W3217
Buffer Solution, PH2 (500ml)	W3161
pH 12.00 Buffer	W3200

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.05].

True Value of CCV2 = 12.00 Control Limits[+/- 0.05].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.2	4.01	09/05/2025	15:10
2	CAL2	1	Water	NA	NA	20.2	7.00	09/05/2025	15:11
3	CAL3	1	Water	NA	NA	20.2	10.02	09/05/2025	15:12
4	ICV	1	Water	NA	NA	20.1	6.99	09/05/2025	15:15
5	CCV1	1	Water	NA	NA	20.2	2.01	09/05/2025	15:20
6	Q3033-01	1	Water	NA	NA	24.3	6.95	09/05/2025	15:30
7	Q3033-01DUP	1	Water	NA	NA	24.6	6.97	09/05/2025	15:33
8	Q3036-03	1	Water	NA	NA	24.7	7.43	09/05/2025	15:40
9	Q3036-03DUP	1	Water	NA	NA	24.9	7.45	09/05/2025	15:44
10	CCV2	1	Water	NA	NA	20.3	12.02	09/05/2025	15:45

WORKLIST(Hardcopy Internal Chain)

137101

WorkList Name : ph w q3033 WorkList ID : 191712 Department : Wet-Chemistry Date : 09-05-2025 14:48:59

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
Q3033-01	BUR-25-0077	Water	pH	Cool 4 deg C	PSEG03	D31	09/05/2025	9040C
Q3036-03	EGR-2	Water	pH	Cool 4 deg C	PSEG03		09/05/2025	9040C

Date/Time 09/05/25 15:00

Raw Sample Received by: SO WOC

Raw Sample Relinquished by: CP SM

Date/Time 09/05/25 17:00

Raw Sample Received by: CP SM

Raw Sample Relinquished by: SO WOC