

Analytical Summary Report

Analysis Method: 365.3
 Parameter: Phosphorus, Total
 Run Number: LB137102

ANALYST: Iwona
 SUPERVISOR REVIEW BY: JIGNESH
 BALANCE ID: WC SC-7

Reagent/Standard	Lot/Log #
calibration std. phosphate 1 ppm	WP114614
calibration std. phosphate 0.5 ppm	WP114613
calibration std. phosphate 0.3 ppm	WP114612
calibration std. phosphate 0.1 ppm	WP114611
calibration std. phosphate 0.05 ppm	WP114610
calibration std. 0 ppm	WP114609
phosphate CCV std.	WP114615
5N sulfuric acid	WP112831
Combined reagent	WP114622
Phenolphthalein indicator	WP113378
Sodium hydroxide, 1N	WP113878
Phosphosphate LOQ std, 0.05PPM	WP114619
Phosphate ICV-LCS Std	WP114616
Phosphate LOD-MDL Std 0.025ppm	WP114620

Intercept: -0.0029 Slope: 0.6521 Regression: 0.999803

Seq	Lab ID	True Value (mg/L)	DF	Initial Volume (mL)	Final Volume (mL)	Absorbance Reading at 880nm	Result (mg/L)	%D	AnalDate	AnalTime
1	CAL1	0.00	1	50	50	0.000	0.004		09/05/2025	12:45
2	CAL2	0.05	1	50	50	0.033	0.055	10	09/05/2025	12:45
3	CAL3	0.10	1	50	50	0.062	0.1	0	09/05/2025	12:46
4	CAL4	0.30	1	50	50	0.183	0.285	-5	09/05/2025	12:46
5	CAL5	0.50	1	50	50	0.325	0.503	0.6	09/05/2025	12:47
6	CAL6	1.00	1	50	50	0.651	1.003	0.3	09/05/2025	12:47

Analytical Summary Report

Analysis Method: 365.3

ANALYST: Iwona

Parameter: Phosphorus, Total

SUPERVISOR REVIEW BY: JIGNESH

Run Number: LB137102

BALANCE ID: WC SC-7

Seq	Lab ID	True Value (mg/l)	DF	Initial Volume (mL)	Final Volume (mL)	Absorbance Reading at 880nm	Result (mg/L)	AnalDate	AnalTime
1	ICV	0.50	1	50	50	0.329	0.509	09/05/2025	12:48
2	ICB		1	50	50	0.003	0.009	09/05/2025	12:48
3	CCV1	0.50	1	50	50	0.324	0.501	09/05/2025	13:05
4	CCB1		1	50	50	0.002	0.008	09/05/2025	13:05
5	RL Check	0.05	1	50	50	0.029	0.049	09/05/2025	13:06
6	PB169586BL		1	1	50	0.003	0.009	09/05/2025	13:06
7	PB169586BS	0.50	1	1	50	0.324	0.501	09/05/2025	13:07
8	Q2893-01		1	1	50	0.015	0.027	09/05/2025	13:07
9	Q2893-02		1	1	50	0.033	0.055	09/05/2025	13:08
10	CCV2	0.50	1	50	50	0.326	0.504	09/05/2025	13:08
11	CCB2		1	50	50	0.001	0.006	09/05/2025	13:09