

Analytical Summary Report

Analysis Method: 1010B

Reviewed By: rubina

Parameter: Flash Point

Supervisor Review By: Iwona

Run Number: LB137108

Ambient Barometric Pressure (mmHg): 765.00

Thermometer ID: Flash Point

Barometric Scale ID: 0511064

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W3194

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	8	28.00	82.4	82.1	09/08/2025	13:35
2	Q3022-01		1	15	100.00	>212.0	>212.0	09/08/2025	14:05
3	Q3022-01DUP		1	15	100.00	>212.0	>212.0	09/08/2025	14:35
4	Q3033-01		1	13	100.00	>212.0	>212.0	09/08/2025	15:05
5	Q3036-03		1	14	100.00	>212.0	>212.0	09/08/2025	15:35

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06

WORKLIST(Hardcopy Internal Chain)

LB137108

WorkList Name : fp-9-08

WorkList ID : 191718

Department : Wet-Chemistry

Date : 09-08-2025 08:21:00

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
Q3022-01	1413	Water	Flash Point	Cool 4 deg C	PSEG03	D31	09/04/2025	1010B
Q3033-01	BUR-25-0077	Water	Flash Point	Cool 4 deg C	PSEG03	D31	09/05/2025	1010B
Q3036-03	EGR-2	Water	Flash Point	Cool 4 deg C	PSEG03		09/05/2025	1010B

Date/Time 09/08/2025 13:25
 Raw Sample Received by: RM CWJ
 Raw Sample Relinquished by: JPLWJ

Date/Time 09/08/2025 16:25
 Raw Sample Received by: JPLWJ
 Raw Sample Relinquished by: RM CWJ