

Analytical Summary Report

Analysis Method: 9045D

Analyst By : JIGNESH

Parameter: Corrosivity

Supervisor Review By : Iwona

Run Number: LB137120

Slope : 98.4

BalanceID: WC SC-7

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W3178
BUFFER PH 7.00 GREEN 1PINT PK6	W3093
PH 10.01 BUFFER,COLOR CD 475ML	W3191
buffer solution pH 7 yellow	W3217
Buffer Solution, PH2 (500ml)	W3161
pH 12.00 Buffer	W3200

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.05].

True Value of CCV2 = 12.00 Control Limits[+/- 0.05].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.2	4.01	09/09/2025	14:10
2	CAL2	1	Water	NA	NA	20.2	7.00	09/09/2025	14:11
3	CAL3	1	Water	NA	NA	20.3	10.02	09/09/2025	14:12
4	ICV	1	Water	NA	NA	20.2	7.00	09/09/2025	14:15
5	CCV1	1	Water	NA	NA	20.2	2.01	09/09/2025	14:20
6	Q3050-01	1	Solid	20.02	20	21.1	8.25	09/09/2025	14:30
7	Q3051-02	1	Solid	20.03	20	20.9	5.65	09/09/2025	14:35
8	Q3054-04	1	Solid	20.04	20	20.7	5.64	09/09/2025	14:47
9	Q3054-08	1	Solid	20.02	20	20.8	8.94	09/09/2025	15:00
10	Q3054-08DUP	1	Solid	20.03	20	20.9	8.95	09/09/2025	15:10
11	CCV2	1	Water	NA	NA	20.2	12.02	09/09/2025	15:11