

Analytical Summary Report

Analysis Method: 180.1

ANALYST: Iwona

Parameter: Turbidity

SUPERVISOR REVIEW BY: JIGNESH

Run Number: LB137139

Reagent/Standard	Lot/Log #
TURBIDITY LOD STD, 0.5NTU	WP114720
Turbidity Calibration std, 0NTU	WP114714
Turbidity Calibration std, 20NTU	WP114716
Turbidity Calibration std, 40NTU	WP114715
Turbidity Calibration std, 1NTU	WP114719
Turbidity Calibration std, 5NTU	WP114718
Turbidity LOQ std, 1.0NTU	WP114721
Turbidity Calibration - CCV std, 10 NTU	WP114717
10 NTU Standard 500 ml	W3116

Intercept: 0.1639

Slope: 0.9595

Regression: 0.9999

Seq	Lab ID	True Value (NTU)	Dilution	Reading	Result (mg/l)	%D	Anal Date	Anal Time
1	CAL1	0	1	0.139	-0.03		09/10/2025	12:15
2	CAL2	1	1	1.050	0.92	-7.6	09/10/2025	12:18
3	CAL3	5	1	4.960	5.00	-0	09/10/2025	12:22
4	CAL4	10	1	9.660	9.90	-1	09/10/2025	12:26
5	CAL5	20	1	19.700	20.36	1.8	09/10/2025	12:30
6	CAL6	40	1	38.400	39.85	-0.4	09/10/2025	12:34



Analytical Summary Report

Analysis Method: 180.1

ANALYST: Iwona

Parameter: Turbidity

SUPERVISOR REVIEW BY: JIGNESH

Run Number: LB137139

Seq	Lab ID	True Value (NTU)	Dilution	Reading	Result (NTU)	AnalDate	Anal Time
1	ICV	10	1	10.300	10.564	09/10/2025	12:37
2	ICB		1	0.144	-0.021	09/10/2025	12:40
3	CCV1	10	1	9.940	10.189	09/10/2025	12:43
4	CCB1		1	0.148	-0.017	09/10/2025	12:46
5	LB137139BL		1	0.147	-0.018	09/10/2025	12:50
6	Q2893-07		1	0.433	0.280	09/10/2025	12:53
7	Q2893-08		1	1.270	1.153	09/10/2025	12:56
8	CCV2	10	1	10.300	10.564	09/10/2025	13:28
9	CCB2		1	0.152	-0.012	09/10/2025	13:32