

Analytical Summary Report

Analysis Method: 1010B

Reviewed By: Iwona

Parameter: Flash Point

Supervisor Review By: JIGNESH

Run Number: LB137164

Ambient Barometric Pressure (mmHg): 760.00

Thermometer ID: Flash Point

Barometric Scale ID: 0511064

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W3194

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	9	28.00	82.4	82.4	09/11/2025	13:45
2	Q3045-02		1	14	100.00	>212.0	>212.0	09/11/2025	14:15
3	Q3045-02DUP		1	14	100.00	>212.0	>212.0	09/11/2025	14:45
4	Q3071-02		1	15	100.00	>212.0	>212.0	09/11/2025	15:15
5	Q3074-02		1	15	100.00	>212.0	>212.0	09/11/2025	15:45

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06

WORKLIST(Hardcopy Internal Chain)

6137164

WorkList Name : fp09-11

WorkList ID : 191810

Department : Wet-Chemistry

Date : 09-11-2025 09:46:47

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
Q3045-02	DNAPL	Water	Flash Point	Cool 4 deg C	PSEG03	D21	09/08/2025	1010B
Q3071-02	4089	Water	Flash Point	Cool 4 deg C	PSEG03	D21	09/10/2025	1010B
Q3074-02	MOO-25-0140	Water	Flash Point	Cool 4 deg C	PSEG03	D31	09/10/2025	1010B

Date/Time 09/11/25 13:40
Raw Sample Received by: 12(wr)
Raw Sample Relinquished by: 28(wr)

Date/Time 09/11/25 16:35
Raw Sample Received by: 28(wr)
Raw Sample Relinquished by: 12(wr)