

Analytical Summary Report

Analysis Method: 1010B

Reviewed By: rubina

Parameter: Flash Point

Supervisor Review By: Iwona

Run Number: LB137191

Ambient Barometric Pressure (mmHg): 765.00

Thermometer ID: Flash Point

Barometric Scale ID: 0511064

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W3194

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	10	28.00	82.4	82.1	09/15/2025	09:30
2	Q3073-01		1	4	20.00	68.0	67.7	09/15/2025	10:00
3	Q3073-01DUP		1	4	20.00	68.0	67.7	09/15/2025	10:30

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06

WORKLIST(Hardcopy Internal Chain)

6137191

WorkList Name : FP-9-12

WorkList ID : 191848

Department : Wet-Chemistry

Date : 09-12-2025 08:23:43

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
Q3073-01	BEL-25-0022	Solid	Flash Point	Cool 4 deg C	PSEG03	D31	09/10/2025	1010B

Date/Time 09/15/2025 08.30
 Raw Sample Received by: RM CWG
 Raw Sample Relinquished by: RM CWG

Date/Time 09/15/2025 10.20
 Raw Sample Received by: RM CWG
 Raw Sample Relinquished by: RM CWG