

Analytical Summary Report

Analysis Method: 1010B

Reviewed By: rubina

Parameter: Flash Point

Supervisor Review By: Iwona

Run Number: LB137192

Ambient Barometric Pressure (mmHg): 765.00

Thermometer ID: Flash Point

Barometric Scale ID: 0511064

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W3194

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	10	28.00	82.4	82.1	09/15/2025	09:30
2	Q3088-02		1	14	100.00	>212.0	>212.0	09/15/2025	13:10
3	Q3088-02DUP		1	14	100.00	>212.0	>212.0	09/15/2025	13:40
4	Q3090-01		1	15	100.00	>212.0	>212.0	09/15/2025	14:10

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06

WORKLIST(Hardcopy Internal Chain)

6137192

WorkList Name : fp-9-15

WorkList ID : 191867

Department : Wet-Chemistry

Date : 09-15-2025 08:15:19

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
Q3088-02	90525-LQ	Water	Flash Point	Cool 4 deg C	PSEG03	D31	09/12/2025	1010B
Q3090-01	90825-B	Water	Flash Point	Cool 4 deg C	PSEG03	D41	09/12/2025	1010B

Date/Time 09/15/2025 08:30
 Raw Sample Received by: RM
 Raw Sample Relinquished by: JP(COC)

Date/Time 09/15/2025 15:40
 Raw Sample Received by: JP(COC)
 Raw Sample Relinquished by: RM