

Analytical Summary Report

Analysis Method: 9040C

Analyst By : JIGNESH

Parameter: pH

Supervisor Review By : Iwona

Run Number: LB137205

Slope : 98.6

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W3178
BUFFER PH 7.00 GREEN 1PINT PK6	W3093
PH 10.01 BUFFER,COLOR CD 475ML	W3191
buffer solution pH 7 yellow	W3217
Buffer Solution, PH2 (500ml)	W3161
pH 12.00 Buffer	W3200

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.05].

True Value of CCV2 = 12.00 Control Limits[+/- 0.05].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.2	4.01	09/16/2025	14:15
2	CAL2	1	Water	NA	NA	20.2	7.01	09/16/2025	14:16
3	CAL3	1	Water	NA	NA	20.2	10.02	09/16/2025	14:17
4	ICV	1	Water	NA	NA	20.3	7.00	09/16/2025	14:19
5	CCV1	1	Water	NA	NA	20.2	2.01	09/16/2025	14:22
6	Q3113-01	1	Water	NA	NA	20.1	8.49	09/16/2025	14:30
7	Q3113-01DUP	1	Water	NA	NA	20.2	8.52	09/16/2025	14:33
8	Q3116-01	1	Water	NA	NA	22.6	7.15	09/16/2025	14:40
9	Q3116-01DUP	1	Water	NA	NA	22.8	7.17	09/16/2025	14:44
10	CCV2	1	Water	NA	NA	20.2	12.02	09/16/2025	14:45

WORKLIST(Hardcopy Internal Chain)

VB 137205

WorkList Name : ph q3116 WorkList ID : 191896 Department : Wet-Chemistry Date : 09-16-2025 13:58:02

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
Q3113-01	TRE-25-0101	Water	pH	Cool 4 deg C	PSEG03	D31	09/16/2025	9040C
Q3116-01	50789	Water	pH	Cool 4 deg C	PSEG03	D21	09/16/2025	9040C

Date/Time 09-16-25 14:10
 Raw Sample Received by: IB WJC
 Raw Sample Relinquished by: CR

Date/Time 09-16-25 16:10
 Raw Sample Received by: CP
 Raw Sample Relinquished by: JB