

## Analytical Summary Report

Analysis Method: 9045D

Analyst By : JIGNESH

Parameter: Corrosivity

Supervisor Review By : Iwona

Run Number: LB137216

Slope : 98.5

BalanceID: WC SC-7

pH Meter ID : WC PH METER-1

| Calibration Standards          | Chemtech Log# |
|--------------------------------|---------------|
| PH 4 BUFFER SOLUTION           | W3178         |
| BUFFER PH 7.00 GREEN 1PINT PK6 | W3093         |
| PH 10.01 BUFFER,COLOR CD 475ML | W3191         |
| buffer solution pH 7 yellow    | W3217         |
| Buffer Solution, PH2 (500ml)   | W3161         |
| pH 12.00 Buffer                | W3200         |

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.05].

True Value of CCV2 = 12.00 Control Limits[+/- 0.05].

| Seq | LabID       | DF | Matrix | Weight (gm) | Volume (ml) | Temperature (°C) | Result (pH) | Anal Date  | Anal Time |
|-----|-------------|----|--------|-------------|-------------|------------------|-------------|------------|-----------|
| 1   | CAL1        | 1  | Water  | NA          | NA          | 20.2             | 4.01        | 09/17/2025 | 13:20     |
| 2   | CAL2        | 1  | Water  | NA          | NA          | 20.2             | 7.00        | 09/17/2025 | 13:21     |
| 3   | CAL3        | 1  | Water  | NA          | NA          | 20.3             | 10.02       | 09/17/2025 | 13:22     |
| 4   | ICV         | 1  | Water  | NA          | NA          | 20.2             | 7.01        | 09/17/2025 | 13:25     |
| 5   | CCV1        | 1  | Water  | NA          | NA          | 20.2             | 2.01        | 09/17/2025 | 13:30     |
| 6   | Q3117-02    | 1  | Solid  | 20.02       | 20          | 22.1             | 10.45       | 09/17/2025 | 13:40     |
| 7   | Q3117-02DUP | 1  | Solid  | 20.04       | 20          | 22.3             | 10.47       | 09/17/2025 | 13:44     |
| 8   | Q3122-01    | 1  | Solid  | 20.03       | 20          | 22.6             | 5.97        | 09/17/2025 | 13:50     |
| 9   | CCV2        | 1  | Water  | NA          | NA          | 20.2             | 12.02       | 09/17/2025 | 14:00     |

9/13/25 16

WORKLIST(Hardcopy Internal Chain)

WorkList Name : corrosivity q31117      WorkList ID : 191923      Department : Wet-Chemistry      Date : 09-17-2025 12:56:44

| Sample   | Customer Sample | Matrix | Test        | Preservative | Customer | Raw Sample Storage Location | Collect Date | Method |
|----------|-----------------|--------|-------------|--------------|----------|-----------------------------|--------------|--------|
| Q3117-02 | 7211986-357     | Solid  | Corrosivity | Cool 4 deg C | PSEG03   | D21                         | 09/16/2025   | 9045D  |
| Q3122-01 | FOTP915         | Solid  | Corrosivity | Cool 4 deg C | PSEG04   | D31                         | 09/15/2025   | 9045D  |

Date/Time 09-17-25 13:10  
Raw Sample Received by: ef wcl  
Raw Sample Relinquished by: R3 CETA-(06)

Date/Time 09-17-25 16:00  
Raw Sample Received by: R3 CETA-(06)  
Raw Sample Relinquished by: JO GDC