

## Analytical Summary Report

Analysis Method: 9040C

Parameter: pH

Run Number: LB137224

Analyst By : gorge

Supervisor Review By : Iwona

Slope : NA

pH Meter ID : Field-Meter

| Calibration Standards          | Chemtech Log# |
|--------------------------------|---------------|
| PH 4 BUFFER SOLUTION           | W3178         |
| buffer solution pH 7 yellow    | W3217         |
| PH 10.01 BUFFER,COLOR CD 475ML | W3191         |
| BUFFER PH 7.00 GREEN 1PINT PK6 | W3093         |

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 7.00 Control Limits[+/- 0.05].

True Value of CCV2 = 7.00 Control Limits[+/- 0.05].

| Seq | LabID       | DF | Matrix | Weight (gm) | Volume (ml) | Temperature (°C) | Result (pH) | Anal Date  | Anal Time |
|-----|-------------|----|--------|-------------|-------------|------------------|-------------|------------|-----------|
| 1   | CAL1        | 1  | Water  | NA          | NA          | 21.1             | 4.02        | 09/18/2025 | 10:05     |
| 2   | CAL2        | 1  | Water  | NA          | NA          | 21.0             | 7.01        | 09/18/2025 | 10:10     |
| 3   | CAL3        | 1  | Water  | NA          | NA          | 21.1             | 10.02       | 09/18/2025 | 10:15     |
| 4   | ICV         | 1  | Water  | NA          | NA          | 21.2             | 7.02        | 09/18/2025 | 10:20     |
| 5   | CCV1        | 1  | Water  | NA          | NA          | 21.0             | 7.00        | 09/18/2025 | 10:25     |
| 6   | Q3019-20    | 1  | Water  | NA          | NA          | 21.3             | 5.82        | 09/18/2025 | 10:35     |
| 7   | Q3019-20DUP | 1  | Water  | NA          | NA          | 21.3             | 5.81        | 09/18/2025 | 10:40     |
| 8   | CCV2        | 1  | Water  | NA          | NA          | 21.2             | 7.01        | 09/18/2025 | 10:45     |