

Analytical Summary Report

Analysis Method: 9040C

Analyst By : JIGNESH

Parameter: pH

Supervisor Review By : Iwona

Run Number: LB137237

Slope : 98.6

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W3178
BUFFER PH 7.00 GREEN 1PINT PK6	W3093
PH 10.01 BUFFER,COLOR CD 475ML	W3191
buffer solution pH 7 yellow	W3217
Buffer Solution, PH2 (500ml)	W3161
pH 12.00 Buffer	W3200

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.05].

True Value of CCV2 = 12.00 Control Limits[+/- 0.05].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.3	4.01	09/19/2025	10:00
2	CAL2	1	Water	NA	NA	20.2	7.00	09/19/2025	10:01
3	CAL3	1	Water	NA	NA	20.3	10.02	09/19/2025	10:05
4	ICV	1	Water	NA	NA	20.2	7.01	09/19/2025	10:10
5	CCV1	1	Water	NA	NA	20.2	2.02	09/19/2025	10:11
6	Q3135-01	1	Water	NA	NA	20.9	7.29	09/19/2025	10:19
7	Q3135-01DUP	1	Water	NA	NA	20.8	7.30	09/19/2025	10:20
8	Q3138-05	1	Water	NA	NA	23.4	6.64	09/19/2025	10:30
9	Q3138-05DUP	1	Water	NA	NA	23.6	6.66	09/19/2025	10:33
10	Q3138-06	1	Water	NA	NA	24.1	7.81	09/19/2025	10:37
11	Q3138-06DUP	1	Water	NA	NA	24.3	7.83	09/19/2025	10:39
12	Q3139-01	1	Water	NA	NA	24.2	7.40	09/19/2025	10:44
13	Q3139-01DUP	1	Water	NA	NA	24.3	7.43	09/19/2025	10:46
14	CCV2	1	Water	NA	NA	20.2	12.02	09/19/2025	10:50

WORKLIST(Hardcopy Internal Chain)

131237

WorkList Name : ph w q3135 WorkList ID : 191950 Department : Wet-Chemistry Date : 09-19-2025 09:39:33

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
Q3135-01	MH-121	Water	pH	Cool 4 deg C	PSEG03	D41	09/18/2025	9040C
Q3138-05	MOO-25-0264	Water	pH	Cool 4 deg C	PSEG03	D31	09/18/2025	9040C
Q3138-06	MOO-25-0242	Water	pH	Cool 4 deg C	PSEG03	D31	09/18/2025	9040C
Q3139-01	LAW-25-0140	Water	pH	Cool 4 deg C	PSEG03	D31	09/18/2025	9040C

Date/Time 09-19-25 09:45
 Raw Sample Received by: [Signature]
 Raw Sample Relinquished by: [Signature]

Date/Time 09-19-25 12:30
 Raw Sample Received by: [Signature]
 Raw Sample Relinquished by: [Signature]