

LB137240


7470A

INSTRUMENT ID: CV1

Linear

 μ Abs.:

37602



A= 0.0000e+000

B= 2.6321e-004

C= -4.6570e-002

Rho= 0.9996529

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	σ/σ1)
0.0	0.000	-0.044	-0.044	9	0.000	9	0				-
0.2	0.200	0.183	-0.017	872	0.0 %	872	0				-9
2.5	2.500	2.523	0.023	9761	0.0 %	9761	0				1
5.0	5.000	5.010	0.010	19213	0.0 %	19213					0
7.5	7.500	7.678	0.178	29346	0.0 %	29346					-2
10.0	10.000	9.851	-0.149	37602	0.0 %	37602					1

LB137240 INSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 19 Sep 2025 14:38:19

Type	Sample ID	Extended ID	Conc.	μ Abs.	Units	Type	Std Conc	Method	Date
S	0.0 - 1	50	-	9	PPB	Std	0.0000	7470A	19 Sep 2025 14:42:13
S	0.2 - 1	50.2	-	872	PPB	Std	0.2000	7470A	19 Sep 2025 14:44:31
S	2.5 - 1	52.5	-	9761	PPB	Std	2.5000	7470A	19 Sep 2025 14:46:48
S	5.0 - 1	55	-	19213	PPB	Std	5.0000	7470A	19 Sep 2025 14:49:05
S	7.5 - 1	57.5	-	29346	PPB	Std	7.5000	7470A	19 Sep 2025 14:54:25
S	10.0 - 1	510	-	37602	PPB	Std	10.0000	7470A	19 Sep 2025 15:06:09
U	ICV39 - 1	ICV39	4.0977	15745	PPB	SMPL	-7470A	-7470A	19 Sep 2025 15:09:10
U	ICB39 - 1	ICB39	-0.0668	-77	PPB	SMPL	-7470A	-7470A	19 Sep 2025 15:11:25
U	CCV17 - 1	CCV17	4.9410	18949	PPB	SMPL	-7470A	-7470A	19 Sep 2025 15:13:43
U	CCB17 - 1	CCB17	-0.0616	-57	PPB	SMPL	-7470A	-7470A	19 Sep 2025 15:16:05
U	CRA - 1	CRA	0.1603	786	PPB	SMPL	-7470A	-7470A	19 Sep 2025 15:18:23
U	HighStd - 1	HighStd	9.1592	34975	PPB	SMPL	-7470A	-7470A	19 Sep 2025 15:20:38
U	ChkStd - 1	ChkStd	6.7095	25668	PPB	SMPL	-7470A	-7470A	19 Sep 2025 15:22:54
U	PB169758BL - 1	PBW	-0.1453	-375	PPB	SMPL	-7470A	-7470A	19 Sep 2025 15:25:20
U	PB169758BS - 1	LCSW	3.4939	13451	PPB	SMPL	-7470A	-7470A	19 Sep 2025 15:27:42
U	Q3135-01 - 1	MH-121	1.5174	5942	PPB	SMPL	-7470A	-7470A	19 Sep 2025 15:32:28
U	Q3135-01DUP - 1	MH-121DUP	1.3526	5316	PPB	SMPL	-7470A	-7470A	19 Sep 2025 15:34:52
U	Q3135-01MS - 1	MH-121MS	2.5081	9706	PPB	SMPL	-7470A	-7470A	19 Sep 2025 15:44:22
U	Q3135-01MSD - 1	MH-121MSD	2.4868	9625	PPB	SMPL	-7470A	-7470A	19 Sep 2025 15:46:40
U	CCV18 - 1	CCV18	5.1465	19730	PPB	SMPL	-7470A	-7470A	19 Sep 2025 15:48:58
U	CCB18 - 1	CCB18	-0.1184	-273	PPB	SMPL	-7470A	-7470A	19 Sep 2025 15:51:16
U	Q3138-05 - 1	MOO-25-0264	0.0498	366	PPB	SMPL	-7470A	-7470A	19 Sep 2025 15:53:34
U	Q3138-06 - 1		0.0319	298	PPB	SMPL	-7470A	-7470A	19 Sep 2025 15:55:49
U	Q3139-01 - 1	MOO-25-0242	-0.0274	73	PPB	SMPL	-7470A	-7470A	19 Sep 2025 15:58:05
U	Q3135-01LX5 - 1		0.2459	1111	PPB	SMPL	-7470A	-7470A	19 Sep 2025 16:05:02
U	Q3135-01A - 1		3.9510	15188	PPB	SMPL	-7470A	-7470A	19 Sep 2025 16:07:21
U	CCV19 - 1	CCV19	5.1637	19795	PPB	SMPL	-7470A	-7470A	19 Sep 2025 16:09:38
U	CCB19 - 1	CCB19	-0.0487	-8	PPB	SMPL	-7470A	-7470A	19 Sep 2025 16:14:33