

Analytical Summary Report

Analysis Method: 9040C

Analyst By : JIGNESH

Parameter: pH

Supervisor Review By : Iwona

Run Number: LB137263

Slope : 98.6

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W3178
BUFFER PH 7.00 GREEN 1PINT PK6	W3093
PH 10.01 BUFFER,COLOR CD 475ML	W3191
buffer solution pH 7 yellow	W3217
Buffer Solution, PH2 (500ml)	W3161
pH 12.00 Buffer	W3200

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.05].

True Value of CCV2 = 12.00 Control Limits[+/- 0.05].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.2	4.01	09/22/2025	13:10
2	CAL2	1	Water	NA	NA	20.2	7.01	09/22/2025	13:11
3	CAL3	1	Water	NA	NA	20.3	10.02	09/22/2025	13:12
4	ICV	1	Water	NA	NA	20.3	7.00	09/22/2025	13:15
5	CCV1	1	Water	NA	NA	20.2	2.02	09/22/2025	14:00
6	Q3151-02	1	Water	NA	NA	20.3	11.40	09/22/2025	14:15
7	Q3151-02DUP	1	Water	NA	NA	20.4	11.42	09/22/2025	14:16
8	CCV2	1	Water	NA	NA	20.3	12.02	09/22/2025	14:20

WORKLIST(Hardcopy Internal Chain)

VF 134263

WorkList Name : ph w q3151 WorkList ID : 192005 Department : Wet-Chemistry Date : 09-22-2025 12:40:13

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
Q3151-02	1402	Water	pH	Cool 4 deg C	PSEG03	D31	09/19/2025	9040C

Date/Time 09/22/25 12:50
 Raw Sample Received by: SS (wcl)
 Raw Sample Relinquished by: JT (cm)

Date/Time 09/22/25 14:40
 Raw Sample Received by: JT (cm)
 Raw Sample Relinquished by: gd arc