

Analytical Summary Report

Analysis Method: 1010B

Reviewed By: rubina

Parameter: Flash Point

Supervisor Review By: Iwona

Run Number: LB137267

Ambient Barometric Pressure (mmHg): 760.00

Thermometer ID: Flash Point

Barometric Scale ID: 0511064

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W3194

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	10	28.00	82.4	82.4	09/22/2025	14:50
2	Q3135-01		1	15	100.00	>212.0	>212.0	09/22/2025	15:20
3	Q3135-01DUP		1	15	100.00	>212.0	>212.0	09/22/2025	15:50
4	Q3151-02		1	12	100.00	>212.0	>212.0	09/22/2025	16:20

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06

WORKLIST(Hardcopy Internal Chain)

66 137267

WorkList Name : fp-09-22

WorkList ID : 191988

Department : Wet-Chemistry

Date : 09-22-2025 08:25:30

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
Q3135-01	MH-121	Water	Flash Point	Cool 4 deg C	PSEG03	D41	09/18/2025	1010B
Q3151-02	1402	Water	Flash Point	Cool 4 deg C	PSEG03	D31	09/19/2025	1010B

Date/Time 09/22/2025 14:40
 Raw Sample Received by: 1214 (WC)
 Raw Sample Relinquished by: [Signature]

Date/Time 09/22/2025 17:00
 Raw Sample Received by: [Signature]
 Raw Sample Relinquished by: [Signature]